

LB133684

245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35950

A= 0.0000e+000

B= 2.7481e-004

C= -8.4750e-003

Rho= 0.9995697

Accept=Accepted

Max. Conc:—

10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1/2
0.0	0.000	0.008	0.008	59	0.000	59					-
0.05	0.050	0.059	0.009	245	0.0 %	245	0				18
0.2	0.200	0.238	0.038	898	0.0 %	898					19
2.5	2.500	2.395	-0.105	8747	0.0 %	8747					-4
5.0	5.000	4.952	-0.048	18051	0.0 %	18051					-1
7.5	7.500	7.727	0.227	28148	0.0 %	28148					3
10.0	10.000	9.871	-0.129	35950	0.0 %	35950					-1

LB133684 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 02 Dec 2024 12:24:16

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	59	-	0.0000	245.1	PPB	02 Dec 2024 13:12:29	S	Std
0.05 - 1	50.05	245	-	0.0500	245.1	PPB	02 Dec 2024 13:17:41	S	Std
0.2 - 1	50.2	898	-	0.2000	245.1	PPB	02 Dec 2024 13:21:20	S	Std
2.5 - 1	52.5	8747	-	2.5000	245.1	PPB	02 Dec 2024 13:23:37	S	Std
5.0 - 1	55	18051	-	5.0000	245.1	PPB	02 Dec 2024 13:25:53	S	Std
7.5 - 1	57.5	28148	-	7.5000	245.1	PPB	02 Dec 2024 13:28:09	S	Std
10.0 - 1	570	35950	-	10.0000	245.1	PPB	02 Dec 2024 13:33:06	S	Std
ICV107 - 1	ICV107	14171	3.8858	-245.1		PPB	02 Dec 2024 13:43:43	U	SMPL
ICB107 - 1	ICB107	28	-0.0008	-245.1		PPB	02 Dec 2024 13:45:58	U	SMPL
CCV69 - 1	CCV69	18560	5.0920	-245.1		PPB	02 Dec 2024 13:48:15	U	SMPL
CCB69 - 1	CCB69	28	-0.0008	-245.1		PPB	02 Dec 2024 13:50:30	U	SMPL
CRA - 1	CRA	195	0.0451	-245.1		PPB	02 Dec 2024 13:56:03	U	SMPL
HighStd - 1	HighStd	35930	9.8654	-245.1		PPB	02 Dec 2024 13:58:18	U	SMPL
ChkStd - 1	ChkStd	25140	6.9002	-245.1		PPB	02 Dec 2024 14:00:33	U	SMPL
PB165329BL - 1	PBW	-253	-0.0780	-245.1		PPB	02 Dec 2024 14:02:52	U	SMPL
PB165329BS - 1	LCSW	14242	3.9053	-245.1		PPB	02 Dec 2024 14:05:10	U	SMPL
P4997-07 - 1	14B-(1-4)-COMP	4582	1.2507	-245.1		PPB	02 Dec 2024 14:07:26	U	SMPL
P5018-07 - 1	14B-(1-4)-COMP	1969	0.5326	-245.1		PPB	02 Dec 2024 14:09:44	U	SMPL
P5018-07DUP - 1	14B-(1-4)-COMPDUP	1860	0.5027	-245.1		PPB	02 Dec 2024 14:12:03	U	SMPL
P5018-07MS - 1	14B-(1-4)-COMPMS	12715	3.4857	-245.1		PPB	02 Dec 2024 14:14:22	U	SMPL
P5018-07MSD - 1	14B-(1-4)-COMPMSD	13193	3.6171	-245.1		PPB	02 Dec 2024 14:16:38	U	SMPL
CCV70 - 1	CCV70	17470	4.7924	-245.1		PPB	02 Dec 2024 14:18:56	U	SMPL
CCB70 - 1	CCB70	36	0.0014	-245.1		PPB	02 Dec 2024 14:21:13	U	SMPL
P5018-07LX5 - 1		306	0.0756	-245.1		PPB	02 Dec 2024 14:23:30	U	SMPL
P5018-07A - 1		14136	3.8762	-245.1		PPB	02 Dec 2024 14:25:45	U	SMPL
CCV71 - 1	CCV71	17794	4.8815	-245.1		PPB	02 Dec 2024 14:28:01	U	SMPL
CCB71 - 1	CCB71	-14	-0.0123	-245.1		PPB	02 Dec 2024 14:30:19	U	SMPL