

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133690

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/02/2024	15:37
2	CAL2	1	Water	NA	NA	20.3	7.00	12/02/2024	15:38
3	CAL3	1	Water	NA	NA	20.3	10.02	12/02/2024	15:40
4	ICV	1	Water	NA	NA	20.3	7.00	12/02/2024	15:45
5	CCV1	1	Water	NA	NA	20.3	2.01	12/02/2024	15:50
6	P5022-01	1	Solid	20.02	20	22.1	6.17	12/02/2024	16:00
7	P5022-01DUP	1	Solid	20.03	20	22.2	6.18	12/02/2024	16:02
8	CCV2	1	Water	NA	NA	20.3	12.02	12/02/2024	16:05

WORKLIST(Hardcopy Internal Chain)

WP 133690

WorkList Name : PH P5022

WorkList ID : 185903

Department : Wet-Chemistry

Date : 12-02-2024 15:27:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5022-01	TAPIAL2-SB02D-13-112424-00.	Solid	pH	Cool 4 deg C	WEST04	L41	11/24/2024	9045D

Date/Time 12/02/24 15:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 12/02/24 17:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]