

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133691

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/02/2024	16:00
2	CAL2	1	Water	NA	NA	20.3	7.00	12/02/2024	16:01
3	CAL3	1	Water	NA	NA	20.3	10.02	12/02/2024	16:05
4	ICV	1	Water	NA	NA	20.3	7.00	12/02/2024	16:10
5	CCV1	1	Water	NA	NA	20.3	2.01	12/02/2024	16:15
6	P5031-01	1	Water	NA	NA	20.6	7.75	12/02/2024	16:20
7	P5031-01DUP	1	Water	NA	NA	20.7	7.76	12/02/2024	16:21
8	CCV2	1	Water	NA	NA	20.3	12.02	12/02/2024	16:22

WORKLIST(Hardcopy Internal Chain)

133691

WorkList Name : PH P5031

WorkList ID : 185904

Department : Wet-Chemistry

Date : 12-02-2024 15:28:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5031-01	286127	Water	pH	Cool 4 deg C	PSEG03	L61	11/27/2024	9040C

Date/Time 12/02/24 15:30
Raw Sample Received by: JA wet
Raw Sample Relinquished by: JA wet

Date/Time 12/02/24 14:30
Raw Sample Received by: CE sm
Raw Sample Relinquished by: JA wet