

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB133700

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/03/2024	08:50
2	CAL2	1	Water	NA	NA	20.3	7.00	12/03/2024	08:51
3	CAL3	1	Water	NA	NA	20.3	10.02	12/03/2024	08:55
4	ICV	1	Water	NA	NA	20.3	7.01	12/03/2024	09:00
5	CCV1	1	Water	NA	NA	20.2	2.01	12/03/2024	09:03
6	P5053-01	1	Water	NA	NA	20.6	7.65	12/03/2024	09:15
7	P5053-01DUP	1	Water	NA	NA	20.7	7.66	12/03/2024	09:16
8	CCV2	1	Water	NA	NA	20.3	12.02	12/03/2024	09:19

WORKLIST(Hardcopy Internal Chain)

133700

WorkList Name : PH P5053

WorkList ID : 185918

Department : Wet-Chemistry

Date : 12-03-2024 08:43:40

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5053-01	A 34740	Water	pH	Cool 4 deg C	PSEG03	M11	12/02/2024	9040C

Date/Time 12/03/24 08:45
Raw Sample Received by: JB WCC
Raw Sample Relinquished by: JMS

Date/Time 12/03/24 10:00
Raw Sample Received by: CP SN
Raw Sample Relinquished by: JB WCC