

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133701

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	12/03/2024	08:45
2	P4995-02		1	13	100.00	>212.0	>212.0	12/03/2024	09:15
3	P4995-02DUP		1	13	100.00	>212.0	>212.0	12/03/2024	09:45
4	P5031-01		1	14	100.00	>212.0	>212.0	12/03/2024	10:15
5	P5053-01		1	11	100.00	>212.0	>212.0	12/03/2024	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66133701

WorkList Name : FP-12-2-

WorkList ID : 185906

Department : Wet-Chemistry

Date : 12-02-2024 17:27:57

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4995-02	001	Water	Flash Point	Cool 4 deg C	ENTA05	L61	11/25/2024	1010B
P5031-01	286127	Water	Flash Point	Cool 4 deg C	PSEG03	L61	11/27/2024	1010B
P5053-01	34740	Water	Flash Point	Cool 4 deg C	PSEG03	M11	12/02/2024	1010B

Date/Time 12/03/2024 08:35
Raw Sample Received by: RIM
Raw Sample Relinquished by: [Signature]

Date/Time 12/03/2024 11:25
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: RIM