

PERCENT SOLID

Supervisor: rubina
Analyst: jignesh
Date: 12/4/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:35
In Date: 12/03/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 12/04/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133713

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5056-01	MBHXX9	1	1.15	8.80	9.95	7.89	76.6	
P5056-02	MBHKY0	2	1.13	8.66	9.79	8.18	81.4	
P5056-03	MBHKY1	3	1.16	8.44	9.6	8.34	85.1	
P5056-04	MBHKY2	4	1.19	8.53	9.72	8.81	89.3	
P5056-05	MBHKY3	5	1.19	8.71	9.9	9.22	92.2	
P5056-06	MBHKY4	6	1.15	8.43	9.58	9.1	94.3	
P5056-07	MBHKY5	7	1.17	8.80	9.97	9.31	92.5	
P5056-08	MBHL35	8	1.15	8.72	9.87	7.89	77.3	
P5056-09	MBHL36	9	1.18	8.67	9.85	7.47	72.5	
P5056-10	MBHL37	10	1.16	8.81	9.97	8.48	83.1	
P5056-11	MBHL38	11	1.14	8.40	9.54	8.32	85.5	
P5056-12	MBHL39	12	1.16	8.50	9.66	8.49	86.2	
P5056-13	MBHL40	13	1.19	8.42	9.61	8.75	89.8	
P5056-14	MBHL41	14	1.12	8.74	9.86	9.15	91.9	
P5056-15	MBHL42	15	1.11	8.86	9.97	6.68	62.9	
P5056-16	MBHL43	16	1.15	8.73	9.88	8.08	79.4	
P5056-17	MBHL44	17	1.18	8.42	9.6	8.00	81.0	
P5056-18	MBHL45	18	1.14	8.74	9.88	8.62	85.6	
P5056-19	MBHL46	19	1.15	8.40	9.55	8.86	91.8	
P5056-20	MBHL47	20	1.19	8.65	9.84	9.35	94.3	
P5056-21	MBHL47D	21	1.19	8.65	9.84	9.35	94.3	
P5056-22	MBHL47S	22	1.19	8.65	9.84	9.35	94.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$