


LB133714

SFAM01.1

INSTRUMENT ID: CV1



A= 0.0000e+000
 B= 2.5128e-004 slope
 C= -8.0939e-002 y-intercept
 Rho= 0.9997324
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/10
0.00	0.000	-0.041	-0.041	158	0.000	158	0				1
0.05	0.050					442	0				5
0.20	0.200	0.190	-0.010	1078	0.0 %	1078					3
2.50	2.500	2.587	0.087	10619	0.0 %	10619					2
5.00	5.000	4.911	-0.089	19865	0.0 %	19865					2
7.50	7.500	7.638	0.138	30719	0.0 %	30719					1
10.0	10.000	9.915	-0.085	39780	0.0 %	39780					

LB133714

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	158	-		0 SFAM01.1	PPB	12/3/2024 15:52	Std
	0.2 S01	1078	-		0.2 SFAM01.1	PPB	12/3/2024 15:54	Std
	2.5 S02	10619	-		2.5 SFAM01.1	PPB	12/3/2024 15:57	Std
	5 S03	19865	-		5 SFAM01.1	PPB	12/3/2024 15:59	Std
	7.5 S04	30719	-		7.5 SFAM01.1	PPB	12/3/2024 16:01	Std
	10 S05	39780	-		10 SFAM01.1	PPB	12/3/2024 16:03	Std
ICV004	ICV004	16550	4.0778 -		SFAM01.1	PPB	12/3/2024 16:06	SMPL
ICB004	ICB004	21	-0.0757 -		SFAM01.1	PPB	12/3/2024 16:09	SMPL
CCV079	CCV079	20200	4.9949 -		SFAM01.1	PPB	12/3/2024 16:11	SMPL
CCB079	CCB079	-2	-0.0814 -		SFAM01.1	PPB	12/3/2024 16:13	SMPL
PB165352BL	PBS352	222	-0.0252 -		SFAM01.1	PPB	12/3/2024 16:15	SMPL
P4917-01	MX1007	182784	45.8491 -		SFAM01.1	PPB	12/3/2024 16:25	SMPL
P4917-01DLX10	MX1007	19608	4.8462 -		SFAM01.1	PPB	12/3/2024 16:35	SMPL
CCV080	CCV080	19042	4.7039 -		SFAM01.1	PPB	12/3/2024 16:37	SMPL
CCB080	CCB080	-58	-0.0955 -		SFAM01.1	PPB	12/3/2024 16:39	SMPL