



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/5/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:55
In Date: 12/04/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:47
Out Date: 12/05/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133723

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5058-01	MBHL80	1	1.15	8.59	9.74	8.56	86.3	
P5058-02	MBHL81	2	1.15	8.56	9.71	8.65	87.6	
P5058-03	MBHL82	3	1.15	8.64	9.79	9.00	90.9	
P5058-04	MBHL82D	4	1.15	8.64	9.79	9.00	90.9	
P5058-05	MBHL82S	5	1.15	8.64	9.79	9.00	90.9	
P5058-06	MBHL83	6	1.17	8.81	9.98	8.94	88.2	
P5058-07	MBHL84	7	1.14	8.57	9.71	7.37	72.7	
P5058-08	MBHL85	8	1.15	8.37	9.52	7.74	78.7	
P5058-09	MBHL86	9	1.16	8.54	9.7	7.97	79.7	
P5058-10	MBHL87	10	1.13	8.67	9.8	7.98	79.0	
P5058-11	MBHL88	11	1.15	8.56	9.71	7.76	77.2	
P5058-12	MBHL89	12	1.15	8.45	9.6	8.45	86.4	
P5058-13	MBHL90	13	1.15	8.83	9.98	8.92	88.0	
P5058-14	MBHL98	14	1.14	8.51	9.65	6.98	68.6	
P5058-15	MBHL99	15	1.13	8.52	9.65	7.71	77.2	
P5058-16	MBHLA0	16	1.15	8.73	9.88	8.18	80.5	
P5058-17	MBHLA1	17	1.15	8.78	9.93	8.7	86.0	
P5058-18	MBHLA2	18	1.13	8.62	9.75	9.04	91.8	
P5058-19	MBHLA3	19	1.13	8.48	9.61	9.11	94.1	
P5058-20	MBHLA4	20	1.14	8.39	9.53	8.98	93.4	
P5058-21	MBHLC6	21	1.14	8.70	9.84	6.96	66.9	
P5058-22	MBHLC7	22	1.14	8.52	9.66	7.47	74.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$