

LB133725


SFAM01.1

INSTRUMENT ID: CVJ

Linear

μ Abs.:

4445E



A= 0.0000e+000

B= 2.2693e-004 *slope*

C= -1.9797e-002 *y-intercept*

Rho= 0.9998776

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>0.0</i>
0.00	0.000	0.024	0.024	191	0.000	191	0				<i>=</i>
0.05	0.050					442	0				<i>4</i>
0.20	0.200	0.208	0.008	1005	0.0 %	1005					<i>2</i>
2.50	2.500	2.544	0.044	11297	0.0 %	11297					<i>-2</i>
5.00	5.000	4.895	-0.105	21656	0.0 %	21656					<i>-1</i>
7.50	7.500	7.460	-0.040	32962	0.0 %	32962					<i>1</i>
10.0	10.000	10.069	0.069	44458	0.0 %	44458					

LB133725
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	191	-		0 SFAM01.1	PPB	12/4/2024 10:18	Std
	0.2 S01	1005	-		0.2 SFAM01.1	PPB	12/4/2024 10:21	Std
	2.5 S02	11297	-		2.5 SFAM01.1	PPB	12/4/2024 10:23	Std
	5 S03	21656	-		5 SFAM01.1	PPB	12/4/2024 10:25	Std
	7.5 S04	32962	-		7.5 SFAM01.1	PPB	12/4/2024 10:27	Std
	10 S05	44458	-		10 SFAM01.1	PPB	12/4/2024 10:30	Std
ICV005	ICV005	17923	4.0475 -		SFAM01.1	PPB	12/4/2024 10:33	SMPL
ICB005	ICB005	87	-0.0001 -		SFAM01.1	PPB	12/4/2024 10:35	SMPL
CCV081	CCV081	21459	4.85 -		SFAM01.1	PPB	12/4/2024 10:37	SMPL
CCB081	CCB081	46	-0.0094 -		SFAM01.1	PPB	12/4/2024 10:40	SMPL
PB165367BL	PBW367	24	-0.0144 -		SFAM01.1	PPB	12/4/2024 10:42	SMPL
P4776-01	MDL-WATER	371	0.0644 -		SFAM01.1	PPB	12/4/2024 10:47	SMPL
PB165368BL	PBW368	27	-0.0137 -		SFAM01.1	PPB	12/4/2024 10:49	SMPL
P4917-05	MX1011	36850	8.3427 -		SFAM01.1	PPB	12/4/2024 10:54	SMPL
CCV082	CCV082	20515	4.6358 -		SFAM01.1	PPB	12/4/2024 10:58	SMPL
CCB082	CCB082	-62	-0.0339 -		SFAM01.1	PPB	12/4/2024 11:01	SMPL