

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB133735

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	12/04/2024	11:40
2	P5070-01		1	11	100.00	>212.0	>212.0	12/04/2024	12:10
3	P5070-01DUP		1	11	100.00	>212.0	>212.0	12/04/2024	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66133735

WorkList Name : fp-s12-04

WorkList ID : 185945

Department : Wet-Chemistry

Date : 12-04-2024 08:24:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5070-01	OIL	Solid	Flash Point	Cool 4 deg C	PSEG03	L41	12/03/2024	1010B

Date/Time 12/04/2024 09:45
Raw Sample Received by: RM CWG
Raw Sample Relinquished by: ADWC

Date/Time 12/04/2024 13:15
Raw Sample Received by: ADWC
Raw Sample Relinquished by: RM CWG