


LB133738

7471B

INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.055	-0.055	80	0.000	80					0.00
0.2	0.200	0.172	-0.028	950	0.0 %	950					-14
2.5	2.500	2.457	-0.043	9708	0.0 %	9708					-2
5.0	5.000	5.176	0.176	20126	0.0 %	20126					4
7.5	7.500	7.646	0.146	29587	0.0 %	29587					2
10.0	10.000	9.804	-0.196	37858	0.0 %	37858					-2

LB133738 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Dec 2024 15:03:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	80	-	0.0000	7471B	PPB	04 Dec 2024 15:05:59	S	Std
0.2 - 1	50.2	950	-	0.2000	7471B	PPB	04 Dec 2024 15:08:15	S	Std
2.5 - 1	52.5	9708	-	2.5000	7471B	PPB	04 Dec 2024 15:10:32	S	Std
5.0 - 1	55	20126	-	5.0000	7471B	PPB	04 Dec 2024 15:12:49	S	Std
7.5 - 1	57.5	29587	-	7.5000	7471B	PPB	04 Dec 2024 15:17:46	S	Std
10.0 - 1	510	37858	-	10.0000	7471B	PPB	04 Dec 2024 15:23:10	S	Std
ICV07 - 1	ICV07	17302	3.9410	-	7471B	PPB	04 Dec 2024 15:28:29	U	SMPL
ICB07 - 1	ICB07	89	-0.0530	-	7471B	PPB	04 Dec 2024 15:38:08	U	SMPL
CCV85 - 1	CCV85	20546	5.2860	-	7471B	PPB	04 Dec 2024 15:40:22	U	SMPL
CCB85 - 1	CCB85	83	-0.0545	-	7471B	PPB	04 Dec 2024 15:45:24	U	SMPL
CRA - 1	CRA	948	0.1712	-	7471B	PPB	04 Dec 2024 15:47:38	U	SMPL
HighStd - 1	HighStd	38667	10.0153	-	7471B	PPB	04 Dec 2024 15:49:54	U	SMPL
ChkStd - 1	ChkStd	27496	7.0999	-	7471B	PPB	04 Dec 2024 15:52:08	U	SMPL
PB165372BL - 1	PBS	349	0.0149	-	7471B	PPB	04 Dec 2024 15:54:30	U	SMPL
PB165372BS - 1	LCSS	18392	4.7238	-	7471B	PPB	04 Dec 2024 15:56:50	U	SMPL
P5052-01 - 1	OR-640-COMP-50	4107	0.9957	-	7471B	PPB	04 Dec 2024 15:59:06	U	SMPL
P5052-07 - 1	OR-640-COMP-51	1812	0.3967	-	7471B	PPB	04 Dec 2024 16:01:25	U	SMPL
P5052-13 - 1	OR-640-COMP-62	2685	0.6245	-	7471B	PPB	04 Dec 2024 16:03:44	U	SMPL
P5052-19 - 1	OR-640-COMP-63	1026	0.1916	-	7471B	PPB	04 Dec 2024 16:06:03	U	SMPL
P5052-19DUP - 1	OR-640-COMP-63DUP	1144	0.2224	-	7471B	PPB	04 Dec 2024 16:11:04	U	SMPL
CCV86 - 1	CCV86	21051	5.4178	-	7471B	PPB	04 Dec 2024 16:13:19	U	SMPL
CCB86 - 1	CCB86	320	0.0073	-	7471B	PPB	04 Dec 2024 16:18:09	U	SMPL
P5052-19MS - 1	OR-640-COMP-63MS	18048	4.6341	-	7471B	PPB	04 Dec 2024 16:20:24	U	SMPL
P5052-19MSD - 1	OR-640-COMP-63MSD	19726	5.0720	-	7471B	PPB	04 Dec 2024 16:22:39	U	SMPL
P5069-01 - 1	A3988-SLUDGE	2470	0.5684	-	7471B	PPB	04 Dec 2024 16:24:58	U	SMPL
P5072-01 - 1	MH-737	1691	0.3651	-	7471B	PPB	04 Dec 2024 16:27:16	U	SMPL
P5072-05 - 1	MH-738	895	0.1574	-	7471B	PPB	04 Dec 2024 16:29:33	U	SMPL
P5052-19LX5 - 1		146	-0.0381	-	7471B	PPB	04 Dec 2024 16:31:50	U	SMPL
P5052-19A - 1		18471	4.7445	-	7471B	PPB	04 Dec 2024 16:34:05	U	SMPL
CCV87 - 1	CCV87	20624	5.3064	-	7471B	PPB	04 Dec 2024 16:38:58	U	SMPL
CCB87 - 1	CCB87	517	0.0587	-	7471B	PPB	04 Dec 2024 16:41:15	U	SMPL