

LB133746


7470A

INSTRUMENT ID: CVI

Linear

 μ Abs.:

40395



A= 0.0000e+000

B= 2.4589e-004

C= -5.9893e-002

Rho= 0.9995698

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.032	-0.032	114	0.000	114	0				0/00
0.2	0.200	0.172	-0.028	943	0.0 %	943					14
2.5	2.500	2.432	-0.068	10134	0.0 %	10134					13
5.0	5.000	5.214	0.214	21450	0.0 %	21450					4
7.5	7.500	7.541	0.041	30912	0.0 %	30912					1
10.0	10.000	9.873	-0.127	40395	0.0 %	40395					1

LB133746 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Dec 2024 07:24:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	114	-	0.0000	7470A	PPB	05 Dec 2024 09:21:24	S	Std
0.2 - 1	50.2	943	-	0.2000	7470A	PPB	05 Dec 2024 09:23:41	S	Std
2.5 - 1	52.5	10134	-	2.5000	7470A	PPB	05 Dec 2024 09:25:58	S	Std
5.0 - 1	55	21450	-	5.0000	7470A	PPB	05 Dec 2024 09:30:59	S	Std
7.5 - 1	57.5	30912	-	7.5000	7470A	PPB	05 Dec 2024 09:35:58	S	Std
10.0 - 1	57.6	40395	-	10.0000	7470A	PPB	05 Dec 2024 09:40:54	S	Std
ICV08 - 1	ICV08	17958	4.3557	-	7470A	PPB	05 Dec 2024 09:46:15	U	SMPL
ICB08 - 1	ICB08	5	-0.0587	-	7470A	PPB	05 Dec 2024 09:51:03	U	SMPL
CCV88 - 1	CCV88	21030	5.1111	-	7470A	PPB	05 Dec 2024 09:53:17	U	SMPL
CCB88 - 1	CCB88	71	-0.0424	-	7470A	PPB	05 Dec 2024 09:55:32	U	SMPL
CRA - 1	CRA	938	0.1707	-	7470A	PPB	05 Dec 2024 09:57:49	U	SMPL
HighStd - 1	HighStd	39423	9.6337	-	7470A	PPB	05 Dec 2024 10:00:04	U	SMPL
ChkStd - 1	ChkStd	27726	6.7575	-	7470A	PPB	05 Dec 2024 10:02:20	U	SMPL
PB165374BL - 1	PBW	197	-0.0115	-	7470A	PPB	05 Dec 2024 10:07:16	U	SMPL
PB165374BS - 1	LCSW	15644	3.7867	-	7470A	PPB	05 Dec 2024 10:09:32	U	SMPL
P5052-02 - 1	OR-640-COMP-50	-78	-0.0791	-	7470A	PPB	05 Dec 2024 10:11:48	U	SMPL
P5052-08 - 1	OR-640-COMP-51	80	-0.0402	-	7470A	PPB	05 Dec 2024 10:14:06	U	SMPL
P5052-14 - 1	OR-640-COMP-62	31	-0.0523	-	7470A	PPB	05 Dec 2024 10:16:23	U	SMPL
P5052-20 - 1	OR-640-COMP-63	100	-0.0353	-	7470A	PPB	05 Dec 2024 10:18:41	U	SMPL
P5069-02 - 1	A3988-SLUDGE	160	-0.0206	-	7470A	PPB	05 Dec 2024 10:20:56	U	SMPL
CCV89 - 1	CCV89	20725	5.0361	-	7470A	PPB	05 Dec 2024 10:23:11	U	SMPL
CCB89 - 1	CCB89	145	-0.0242	-	7470A	PPB	05 Dec 2024 10:28:02	U	SMPL
P5072-04 - 1	MH-737	15	-0.0562	-	7470A	PPB	05 Dec 2024 10:30:16	U	SMPL
P5072-08 - 1	MH-738	41	-0.0498	-	7470A	PPB	05 Dec 2024 10:32:31	U	SMPL
P5072-08DUP - 1	MH-738DUP	75	-0.0415	-	7470A	PPB	05 Dec 2024 10:34:47	U	SMPL
P5072-08MS - 1	MH-738MS	17580	4.2628	-	7470A	PPB	05 Dec 2024 10:37:02	U	SMPL
P5072-08MSD - 1	MH-738MSD	15248	3.6894	-	7470A	PPB	05 Dec 2024 10:39:18	U	SMPL
PB165341TB - 1	PB165341TB	-82	-0.0801	-	7470A	PPB	05 Dec 2024 10:41:36	U	SMPL
P5072-08X5 - 1		-6	-0.0614	-	7470A	PPB	05 Dec 2024 10:43:54	U	SMPL
P5072-08A - 1		16823	4.0766	-	7470A	PPB	05 Dec 2024 10:46:11	U	SMPL
CCB90 - 1	CCB90	20718	5.0344	-	7470A	PPB	05 Dec 2024 10:48:27	U	SMPL
CCB90 - 1	CCB90	-51	-0.0724	-	7470A	PPB	05 Dec 2024 10:50:44	U	SMPL