

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB133748

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/05/2024	08:45
2	CAL2	1	Water	NA	NA	20.2	7.01	12/05/2024	08:46
3	CAL3	1	Water	NA	NA	20.3	10.02	12/05/2024	08:48
4	ICV	1	Water	NA	NA	20.3	7.00	12/05/2024	08:50
5	CCV1	1	Water	NA	NA	20.3	2.01	12/05/2024	08:55
6	P5095-04	1	Solid	20.02	20	20.5	9.57	12/05/2024	09:00
7	P5095-04DUP	1	Solid	20.03	20	20.6	9.58	12/05/2024	09:01
8	P5096-04	1	Solid	20.02	20	20.4	8.57	12/05/2024	09:05
9	P5096-08	1	Solid	20.03	20	20.6	8.71	12/05/2024	09:10
10	P5100-04	1	Solid	20.02	20	20.9	8.67	12/05/2024	09:15
11	P5103-02	1	Solid	20.02	20	20.7	8.64	12/05/2024	09:20
12	P5110-01	1	Solid	20.02	20	20.7	6.78	12/05/2024	09:25
13	P5110-02	1	Solid	20.01	20	20.9	6.32	12/05/2024	09:28
14	CCV2	1	Water	NA	NA	20.3	12.02	12/05/2024	09:30

WORKLIST(Hardcopy Internal Chain)

133748

WorkList Name : corrosivity p5096

WorkList ID : 185990

Department : Wet-Chemistry

Date : 12-05-2024 08:38:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5095-04	MH-764	Solid	Corrosivity	Cool 4 deg C	PSEG03	L11	12/04/2024	9045D
P5096-04	MH-B	Solid	Corrosivity	Cool 4 deg C	PSEG03	L51	12/04/2024	9045D
P5096-08	MH-A	Solid	Corrosivity	Cool 4 deg C	PSEG03	L51	12/04/2024	9045D
P5100-04	3167	Solid	Corrosivity	Cool 4 deg C	PSEG03	L61	12/04/2024	9045D
P5103-02	423	Solid	Corrosivity	Cool 4 deg C	PSEG03	L51	12/04/2024	9045D
P5110-01	ELIZ-COMP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	12/04/2024	9045D
P5110-02	ELIZ-COMP-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	12/04/2024	9045D

Date/Time 12/05/24 08:40

Raw Sample Received by: 18 WLC

Raw Sample Relinquished by: 18 WLC

Date/Time 12/05/24 12:00

Raw Sample Received by: CE

Raw Sample Relinquished by: 18 WLC