

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB133749

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/05/2024	08:45
2	CAL2	1	Water	NA	NA	20.2	7.01	12/05/2024	08:46
3	CAL3	1	Water	NA	NA	20.3	10.02	12/05/2024	08:48
4	ICV	1	Water	NA	NA	20.3	7.00	12/05/2024	08:50
5	CCV1	1	Water	NA	NA	20.1	2.02	12/05/2024	09:33
6	P5070-03	1	Water	NA	NA	20.1	5.72	12/05/2024	09:37
7	P5100-03	1	Water	NA	NA	20.5	6.42	12/05/2024	09:41
8	P5103-01	1	Water	NA	NA	20.4	7.91	12/05/2024	09:47
9	P5106-01	1	Water	NA	NA	20.6	7.69	12/05/2024	09:50
10	P5106-01DUP	1	Water	NA	NA	20.7	7.70	12/05/2024	09:51
11	CCV2	1	Water	NA	NA	20.3	12.02	12/05/2024	09:55

WORKLIST(Hardcopy Internal Chain)

13349

WorkList Name : ph p5106 WorkList ID : 185991 Department : Wet-Chemistry Date : 12-05-2024 08:39:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5070-03	A OIL-WATER	Water	pH	Cool 4 deg C	PSEG03	L61	12/03/2024	9040C
P5100-03	A 3154	Water	pH	Cool 4 deg C	PSEG03	L61	12/04/2024	9040C
P5103-01	A 422	Water	pH	Cool 4 deg C	PSEG03	L51	12/04/2024	9040C
P5106-01	A NWB-2069	Water	pH	Cool 4 deg C	PSEG03	L51	12/04/2024	9040C

Date/Time 12/05/24 08:42
Raw Sample Received by: JBWC
Raw Sample Relinquished by: CFS

Date/Time 12/05/24 12:30
Raw Sample Received by: CFS
Raw Sample Relinquished by: JBWC