

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133750

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/05/2024	08:45
2	CAL2	1	Water	NA	NA	20.2	7.01	12/05/2024	08:46
3	CAL3	1	Water	NA	NA	20.3	10.02	12/05/2024	08:48
4	ICV	1	Water	NA	NA	20.3	7.00	12/05/2024	08:50
5	CCV1	1	Water	NA	NA	20.3	2.02	12/05/2024	10:00
6	P5076-01	1	Solid	20.02	20	20.6	5.27	12/05/2024	10:10
7	P5076-01DUP	1	Solid	20.03	20	20.7	5.28	12/05/2024	10:11
8	CCV2	1	Water	NA	NA	20.3	12.02	12/05/2024	10:15

WORKLIST(Hardcopy Internal Chain)

133750

WorkList Name : ph p5076 WorkList ID : 185992 Department : Wet-Chemistry Date : 12-05-2024 08:40:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5076-01	A-TAPIAL2-SB02I-7.5-120224-00-	Solid	pH	Cool 4 deg C	WEST04	L61	12/02/2024	9045D

Date/Time 2105124 08:47
Raw Sample Received by: JB CDC
Raw Sample Relinquished by: JSC

Date/Time 1205124 13:00
Raw Sample Received by: JSC
Raw Sample Relinquished by: JB CDC