



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/6/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:30
In Date: 12/05/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:42
Out Date: 12/06/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133753

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5079-01	MBHLE9	1	1.15	8.64	9.79	8.36	83.4	
P5079-02	MBHLF0	2	1.15	8.65	9.8	8.88	89.4	
P5079-03	MBHLF1	3	1.14	8.50	9.64	8.77	89.8	
P5079-04	MBHLF2	4	1.14	8.39	9.53	8.48	87.5	
P5079-05	MBHLF3	5	1.14	8.84	9.98	7.62	73.3	
P5079-06	MBHLF4	6	1.13	8.43	9.56	8.62	88.8	
P5079-07	MBHLF5	7	1.15	8.48	9.63	8.16	82.7	
P5079-08	MBHLF6	8	1.17	8.44	9.61	7.75	78.0	
P5079-09	MBHLF7	9	1.15	8.52	9.67	7.89	79.1	
P5079-10	MBHLF8	10	1.15	8.42	9.57	8.39	86.0	
P5079-11	MBHLF8D	11	1.15	8.42	9.57	8.39	86.0	
P5079-12	MBHLF8S	12	1.15	8.42	9.57	8.39	86.0	
P5079-13	MBHLF9	13	1.14	8.50	9.64	8.62	88.0	
P5079-14	MBHLG7	14	1.17	8.63	9.8	7.5	73.3	
P5079-15	MBHLG8	15	1.17	8.73	9.9	7.72	75.0	
P5079-16	MBHLG9	16	1.15	8.82	9.97	8.24	80.4	
P5079-17	MBHLH0	17	1.14	8.51	9.65	8.4	85.3	
P5079-18	MBHLH1	18	1.15	8.72	9.87	8.8	87.7	
P5079-19	MBHLH2	19	1.17	8.39	9.56	8.83	91.3	
P5079-20	MBHLH3	20	1.17	8.46	9.63	8.98	92.3	
P5079-21	MBHLH4	21	1.14	8.84	9.98	7.76	74.9	
P5079-22	MBHLH5	22	1.17	8.80	9.97	7.93	76.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$