

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133758

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	12/06/2024	09:15
2	P5100-04		1	11	100.00	>212.0	>212.0	12/06/2024	09:45
3	P5100-04DUP		1	11	100.00	>212.0	>212.0	12/06/2024	10:15
4	P5103-02		1	4	7.00	44.6	44.3	12/06/2024	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6133758

WorkList Name : fp-s-12-05 WorkList ID : 185998 Department : Wet-Chemistry Date : 12-05-2024 09:45:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5100-04	3167	Solid	Flash Point	Cool 4 deg C	PSEG03	L61	12/04/2024	1010B
P5103-02	423	Solid	Flash Point	Cool 4 deg C	PSEG03	L51	12/04/2024	1010B

Date/Time 12/06/2024 09:05
Raw Sample Received by: RM CWG
Raw Sample Relinquished by: pp/coc

Date/Time 12/06/2024 11:20
Raw Sample Received by: pp/coc
Raw Sample Relinquished by: RM CWG