

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133759

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	12/05/2024	12:45
2	P5100-03		1	12	100.00	>212.0	>212.0	12/05/2024	13:15
3	P5100-03DUP		1	12	100.00	>212.0	>212.0	12/05/2024	13:45
4	P5103-01		1	14	100.00	>212.0	>212.0	12/05/2024	14:15
5	P5106-01		1	11	100.00	>212.0	>212.0	12/05/2024	14:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16133759

WorkList Name : fp-12-04

WorkList ID : 185980

Department : Wet-Chemistry

Date : 12-04-2024 17:06:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5100-03	3154	Water	Flash Point	Cool 4 deg C	PSEG03	L61	12/04/2024	1010B
P5103-01	422	Water	Flash Point	Cool 4 deg C	PSEG03	L51	12/04/2024	1010B
P5106-01	NWB-2069	Water	Flash Point	Cool 4 deg C	PSEG03	L51	12/04/2024	1010B

Date/Time 12/05/2024 12.35
 Raw Sample Received by: RY CWG
 Raw Sample Relinquished by: RY CWG

Date/Time 12/05/2024 15.20
 Raw Sample Received by: RY CWG
 Raw Sample Relinquished by: RY CWG