


L3133761

SFAM01.1

Linear

μ Abs.:

37403



A= 0.0000e+000

B= 2.6912e-004 *slope*

C= -3.9785e-002 *y-intercept*

Rho= 0.9999408

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>%D</i>
0.00	0.000	-0.008	-0.008	118	0.000	118	0				
0.05	0.050					442	0				
0.20	0.200	0.181	-0.019	822	0.0 %	822					
2.50	2.500	2.576	0.076	9718	0.0 %	9718					
5.00	5.000	4.957	-0.043	18568	0.0 %	18568					
7.50	7.500	7.468	-0.032	27896	0.0 %	27896					
10.0	10.000	10.026	0.026	37403	0.0 %	37403					

010D
 =
 -10
 3
 -1
 0
 0

LB133761

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0		118	-		0 SFAM01.1	PPB	12/5/2024 13:21	Std
0.2		822	-		0.2 SFAM01.1	PPB	12/5/2024 13:23	Std
2.5		9718	-		2.5 SFAM01.1	PPB	12/5/2024 13:26	Std
5		18568	-		5 SFAM01.1	PPB	12/5/2024 13:28	Std
7.5		27896	-		7.5 SFAM01.1	PPB	12/5/2024 13:30	Std
10		37403	-		10 SFAM01.1	PPB	12/5/2024 13:35	Std
ICV009	ICV009	14915	3.9742 -		SFAM01.1	PPB	12/5/2024 13:38	SMPL
ICB009	ICB009	152	0.0011 -		SFAM01.1	PPB	12/5/2024 13:43	SMPL
CCV091	CCV091	17710	4.7264 -		SFAM01.1	PPB	12/5/2024 13:45	SMPL
CCB091	CCB091	153	0.0014 -		SFAM01.1	PPB	12/5/2024 13:50	SMPL
PB165405BL	PBW405	118	-0.008 -		SFAM01.1	PPB	12/5/2024 13:53	SMPL
P4776-02	MDL-WATER	378	0.0619 -		SFAM01.1	PPB	12/5/2024 13:57	SMPL
CCV092	CCV092	17770	4.7425 -		SFAM01.1	PPB	12/5/2024 14:00	SMPL
CCB092	CCB092	109	-0.0105 -		SFAM01.1	PPB	12/5/2024 14:04	SMPL