



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/6/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 15:55
In Date: 12/05/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 12/06/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133764

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5086-01	MBHLM6	1	1.15	8.83	9.98	7.36	70.3	
P5086-02	MBHLM7	2	1.15	8.81	9.96	8.08	78.7	
P5086-03	MBHLM8	3	1.16	8.78	9.94	7.89	76.7	
P5086-04	MBHLM9	4	1.16	8.44	9.6	7.45	74.5	
P5086-05	MBHLN0	5	1.15	8.67	9.82	8.34	82.9	
P5086-06	MBHLN1	6	1.15	8.78	9.93	8.56	84.4	
P5086-07	MBHLN2	7	1.18	8.65	9.83	8.3	82.3	
P5086-08	MBHLN2D	8	1.18	8.65	9.83	8.3	82.3	
P5086-09	MBHLN2S	9	1.18	8.65	9.83	8.3	82.3	
P5086-10	MBHLN3	10	1.16	8.81	9.97	7.23	68.9	
P5086-11	MBHLN4	11	1.17	8.47	9.64	7.37	73.2	
P5086-12	MBHLN5	12	1.16	8.58	9.74	8.08	80.7	
P5086-13	MBHLN6	13	1.16	8.82	9.98	8.48	83.0	
P5086-14	MBHLN7	14	1.17	8.45	9.62	8.39	85.4	
P5086-15	MBHLN8	15	1.16	8.58	9.74	8.55	86.1	
P5086-16	MBHLN9	16	1.18	8.64	9.82	8.67	86.7	
P5086-17	MBHLP7	17	1.17	8.65	9.82	7.4	72.0	
P5086-18	MBHLP8	18	1.16	8.65	9.81	7.43	72.5	
P5086-19	MBHLP9	19	1.16	8.57	9.73	7.64	75.6	
P5086-20	MBHLQ0	20	1.15	8.48	9.63	7.94	80.1	
P5086-21	MBHLQ1	21	1.16	8.45	9.61	8.15	82.7	
P5086-22	MBHLQ2	22	1.14	8.69	9.83	8.72	87.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$