

LB133766


SFAM01.1

INSTRUMENT ID: CVJ

Linear

μ Abs.:

39160



A= 0.0000e+000

B= 2.5529e-004 slope

C= -3.2361e-002 y-intercept

Rho= 0.9996148

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	std
0.00	0.000	-0.006	-0.006	103	0.000	103	0				0.00
0.05	0.050					442	0				0.05
0.20	0.200	0.229	0.029	1023	0.0 %	1023					0.20
2.50	2.500	2.557	0.057	10144	0.0 %	10144					2.50
5.00	5.000	4.809	-0.191	18966	0.0 %	18966					5.00
7.50	7.500	7.646	0.146	30076	0.0 %	30076					7.50
10.0	10.000	9.965	-0.035	39160	0.0 %	39160					10.0

LB133766

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	103	-		0 SFAM01.1	PPB	12/5/2024 15:12	Std
	0.2 S01	1023	-		0.2 SFAM01.1	PPB	12/5/2024 15:15	Std
	2.5 S02	10144	-		2.5 SFAM01.1	PPB	12/5/2024 15:17	Std
	5 S03	18966	-		5 SFAM01.1	PPB	12/5/2024 15:19	Std
	7.5 S04	30076	-		7.5 SFAM01.1	PPB	12/5/2024 15:21	Std
	10 S05	39160	-		10 SFAM01.1	PPB	12/5/2024 15:24	Std
ICV010	ICV010	15852	4.0145 -		SFAM01.1	PPB	12/5/2024 15:27	SMPL
ICB010	ICB010	54	-0.0186 -		SFAM01.1	PPB	12/5/2024 15:29	SMPL
CCV093	CCV093	19343	4.9057 -		SFAM01.1	PPB	12/5/2024 15:31	SMPL
CCB093	CCB093	124	-0.0007 -		SFAM01.1	PPB	12/5/2024 15:36	SMPL
PB165413BL	PBS413	211	0.0215 -		SFAM01.1	PPB	12/5/2024 15:38	SMPL
P4776-04	MDL-SOIL	365	0.0608 -		SFAM01.1	PPB	12/5/2024 15:43	SMPL
CCV094	CCV094	19137	4.8531 -		SFAM01.1	PPB	12/5/2024 15:45	SMPL
CCB094	CCB094	214	0.0223 -		SFAM01.1	PPB	12/5/2024 15:50	SMPL