

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB133775

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/06/2024	08:55
2	CAL2	1	Water	NA	NA	20.2	7.00	12/06/2024	08:56
3	CAL3	1	Water	NA	NA	20.3	10.02	12/06/2024	08:58
4	ICV	1	Water	NA	NA	20.3	7.00	12/06/2024	09:00
5	CCV1	1	Water	NA	NA	20.3	2.01	12/06/2024	09:05
6	P5145-01	1	Water	NA	NA	20.1	7.17	12/06/2024	09:10
7	P5145-01DUP	1	Water	NA	NA	20.2	7.18	12/06/2024	09:11
8	CCV2	1	Water	NA	NA	20.3	12.02	12/06/2024	09:15

WORKLIST(Hardcopy Internal Chain)

133775

WorkList Name : PH P5145

WorkList ID : 186056

Department : Wet-Chemistry

Date : 12-06-2024 08:45:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5145-01	286085	Water	pH	Cool 4 deg C	PSEG03	L51	12/05/2024	9040C

Date/Time 12/06/24 08:50

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: