



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/9/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:20
In Date: 12/06/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:31
Out Date: 12/07/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133781

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5089-01	MBHLZ4	1	1.18	8.44	9.62	7.92	79.9	
P5089-02	MBHLZ4D	2	1.18	8.44	9.62	7.92	79.9	
P5089-03	MBHLZ4S	3	1.18	8.44	9.62	7.92	79.9	
P5089-04	MBHLZ5	4	1.18	8.47	9.65	8.3	84.1	
P5089-05	MBHLZ6	5	1.16	8.56	9.72	8.68	87.9	
P5089-06	MBHLZ7	6	1.17	8.41	9.58	8.75	90.1	
P5089-07	MBHLZ8	7	1.19	8.70	9.89	9.06	90.5	
P5089-08	MBHLZ9	8	1.18	8.55	9.73	9.12	92.9	
P5089-09	MBHM00	9	1.16	8.57	9.73	8.73	88.3	
P5089-10	MBHM01	10	1.16	8.77	9.93	7.53	72.6	
P5089-11	MBHM02	11	1.18	8.55	9.73	8.09	80.8	
P5089-12	MBHM03	12	1.19	8.47	9.66	8.28	83.7	
P5089-13	MBHM04	13	1.16	8.55	9.71	7.87	78.5	
P5089-14	MBHM05	14	1.18	8.79	9.97	8.04	78.0	
P5089-15	MBHM06	15	1.17	8.69	9.86	7.76	75.8	
P5089-16	MBHM07	16	1.16	8.70	9.86	8.77	87.5	
P5089-17	MBHM08	17	1.18	8.62	9.8	8.19	81.3	
P5089-18	MBHM09	18	1.19	8.39	9.58	8.38	85.7	
P5089-19	MBHM10	19	1.18	8.56	9.74	8.7	87.9	
P5089-20	MBHM11	20	1.19	8.47	9.66	8.61	87.6	
P5089-21	MBHM12	21	1.18	8.41	9.59	8.87	91.4	
P5089-22	MBHM13	22	1.19	8.42	9.61	8.06	81.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$