



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/9/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:50
In Date: 12/06/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:37
Out Date: 12/07/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133783

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5091-01	MBHLR4	1	1.15	8.68	9.83	9.2	92.7	
P5091-02	MBHLR5	2	1.16	8.71	9.87	9.19	92.2	
P5091-03	MBHLR6	3	1.15	8.44	9.59	8.84	91.1	
P5091-04	MBHLR7	4	1.13	8.44	9.57	8.83	91.2	
P5091-05	MBHLR8	5	1.13	8.74	9.87	7.25	70.0	
P5091-06	MBHLR9	6	1.14	8.83	9.97	7.66	73.8	
P5091-07	MBHLS0	7	1.18	8.56	9.74	8.08	80.6	
P5091-08	MBHLS0D	8	1.18	8.56	9.74	8.08	80.6	
P5091-09	MBHLS0S	9	1.18	8.56	9.74	8.08	80.6	
P5091-10	MBHLS1	10	1.16	8.35	9.51	8.15	83.7	
P5091-11	MBHLS2	11	1.15	8.62	9.77	8.92	90.1	
P5091-12	MBHLS3	12	1.17	8.36	9.53	8.92	92.7	
P5091-13	MBHLS4	13	1.15	8.42	9.57	8.7	89.7	
P5091-14	MBHLS5	14	1.16	8.68	9.84	7.43	72.2	
P5091-15	MBHLS6	15	1.17	8.57	9.74	7.85	77.9	
P5091-16	MBHLS7	16	1.15	8.55	9.7	8.47	85.6	
P5091-17	MBHLS8	17	1.16	8.52	9.68	8.72	88.7	
P5091-18	MBHLS9	18	1.18	8.78	9.96	9.4	93.6	
P5091-19	MBHLT0	19	1.16	8.38	9.54	9.00	93.6	
P5091-20	MBHLT1	20	1.17	8.78	9.95	9.31	92.7	
P5091-21	MBHLT2	21	1.17	8.66	9.83	8.05	79.4	
P5091-22	MBHLT3	22	1.15	8.56	9.71	8.08	81.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$