

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133802

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/07/2024	08:02
2	CAL2	1	Water	NA	NA	20.3	7.00	12/07/2024	08:03
3	CAL3	1	Water	NA	NA	20.3	10.02	12/07/2024	08:05
4	ICV	1	Water	NA	NA	20.3	7.00	12/07/2024	08:10
5	CCV1	1	Water	NA	NA	20.3	2.01	12/07/2024	08:11
6	P5159-02	1	Solid	20.02	20	20.9	6.92	12/07/2024	08:20
7	P5159-02DUP	1	Solid	20.03	20	20.8	6.93	12/07/2024	08:21
8	P5159-03	1	Solid	20.04	20	20.8	7.06	12/07/2024	08:25
9	CCV2	1	Water	NA	NA	20.3	12.02	12/07/2024	08:30

WORKLIST(Hardcopy Internal Chain)

VB 133802

WorkList Name : PH P5159 S

WorkList ID : 186083

Department : Wet-Chemistry

Date : 12-07-2024 07:51:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5159-02	✓ 102824-A	Solid	pH	Cool 4 deg C	PSEG03	L61	12/06/2024	9045D
P5159-03	✓ 102824-B	Solid	pH	Cool 4 deg C	PSEG03	L61	12/06/2024	9045D

Date/Time 12/07/24 08:00

Raw Sample Received by: JH (WS)

Raw Sample Relinquished by: RM (WS)

Date/Time 12/07/24 13:30

Raw Sample Received by: RM (WS)

Raw Sample Relinquished by: JH (WS)