

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133803

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/07/2024	08:02
2	CAL2	1	Water	NA	NA	20.3	7.00	12/07/2024	08:03
3	CAL3	1	Water	NA	NA	20.3	10.02	12/07/2024	08:05
4	ICV	1	Water	NA	NA	20.3	7.00	12/07/2024	08:10
5	CCV1	1	Water	NA	NA	20.3	2.01	12/07/2024	08:35
6	P5160-01	1	Water	NA	NA	20.1	5.18	12/07/2024	08:40
7	P5160-01DUP	1	Water	NA	NA	20.2	5.19	12/07/2024	08:41
8	P5170-02	1	Water	NA	NA	20.6	6.13	12/07/2024	08:47
9	P5170-03	1	Water	NA	NA	20.4	6.07	12/07/2024	08:50
10	CCV2	1	Water	NA	NA	20.3	12.02	12/07/2024	09:00

WORKLIST(Hardcopy Internal Chain)

WB 133803

WorkList Name : PH P5170

WorkList ID : 186084

Department : Wet-Chemistry

Date : 12-07-2024 07:52:01

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5160-01	112524	Water	pH	Cool 4 deg C	PSEG03	L61	12/06/2024	9040C
P5170-02	120324-A	Water	pH	Cool 4 deg C	PSEG03	L41	12/06/2024	9040C
P5170-03	120324-B	Water	pH	Cool 4 deg C	PSEG03	L41	12/06/2024	9040C

Date/Time 12/07/24 08:00

Raw Sample Received by: RM CWG

Raw Sample Relinquished by: RM CWG

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: