

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB133804

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/07/2024	08:02
2	CAL2	1	Water	NA	NA	20.3	7.00	12/07/2024	08:03
3	CAL3	1	Water	NA	NA	20.3	10.02	12/07/2024	08:05
4	ICV	1	Water	NA	NA	20.3	7.00	12/07/2024	08:10
5	CCV1	1	Water	NA	NA	20.1	2.01	12/07/2024	09:05
6	P5174-02	1	Solid	20.02	20	20.5	6.88	12/07/2024	09:15
7	P5196-04	1	Solid	20.03	20	20.7	6.82	12/07/2024	09:20
8	P5196-04DUP	1	Solid	20.04	20	20.8	6.83	12/07/2024	09:21
9	CCV2	1	Water	NA	NA	20.3	12.02	12/07/2024	09:25

WORKLIST(Hardcopy Internal Chain)

UB 133804

WorkList Name : corrosivity p5196

WorkList ID : 186082

Department : Wet-Chemistry

Date : 12-07-2024 07:50:07

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5174-02	ROLL-OFF-COMP	Solid	Corrosivity	Cool 4 deg C	PSEG03	L51	12/06/2024	9045D
P5196-04	MH-761	Solid	Corrosivity	Cool 4 deg C	PSEG03	L51	12/06/2024	9045D

Date/Time

12/07/24 08:00

Raw Sample Received by:

RM CWC

Raw Sample Relinquished by:

RM CWC

Date/Time

12/07/24

Raw Sample Received by:

RM CWC

Raw Sample Relinquished by:

RM CWC