



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/9/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 14:25
In Date: 12/07/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:49
Out Date: 12/08/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133815

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5130-01	MBHMF4	1	1.15	8.52	9.67	7.33	72.5	
P5130-02	MBHMF5	2	1.15	8.45	9.6	7.48	74.9	
P5130-03	MBHMF6	3	1.14	8.80	9.94	7.65	74.0	
P5130-04	MBHMF7	4	1.2	8.65	9.85	7.81	76.4	
P5130-05	MBHMF8	5	1.15	8.75	9.9	8.19	80.5	
P5130-06	MBHMF9	6	1.19	8.41	9.6	7.89	79.7	
P5130-07	MBHMG3	7	1.15	8.38	9.53	7.28	73.2	
P5130-08	MBHMG4	8	1.15	8.81	9.96	8.04	78.2	
P5130-09	MBHMG5	9	1.15	8.40	9.55	8.07	82.4	
P5130-10	MBHMG6	10	1.15	8.57	9.72	8.28	83.2	
P5130-11	MBHMG7	11	1.16	8.66	9.82	8.56	85.5	
P5130-12	MBHMH2	12	1.15	8.47	9.62	8.47	86.4	
P5130-13	MBHMH3	13	1.17	8.71	9.88	8.85	88.2	
P5130-14	MBHMH4	14	1.17	8.45	9.62	8.51	86.9	
P5130-15	MBHMH5	15	1.16	8.42	9.58	8.48	86.9	
P5130-16	MBHMH5D	16	1.16	8.42	9.58	8.48	86.9	
P5130-17	MBHMH5S	17	1.16	8.42	9.58	8.48	86.9	
P5130-18	MBHML5	18	1.16	8.37	9.53	8.08	82.7	
P5130-19	MBHML6	19	1.17	8.49	9.66	8.58	87.3	
P5130-20	MBHML7	20	1.16	8.44	9.6	8.53	87.3	
P5130-21	MBHML8	21	1.16	8.68	9.84	7.53	73.4	
P5130-22	MBHML9	22	1.17	8.48	9.65	8.11	81.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$