



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/9/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 15:25
In Date: 12/07/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 12/08/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133817

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5071-01	MBHXX5	1	1.15	8.37	9.52	8.83	91.8	
P5071-02	MBHXX5D	2	1.15	8.37	9.52	8.83	91.8	
P5071-03	MBHXX5S	3	1.15	8.37	9.52	8.83	91.8	
P5132-01	MBHME1	4	1.12	8.70	9.82	8.61	86.1	
P5132-02	MBHME2	5	1.15	8.47	9.62	8.26	83.9	
P5132-03	MBHME3	6	1.15	8.43	9.58	8.32	85.1	
P5132-04	MBHMG0	7	1.15	8.45	9.6	7.72	77.8	
P5132-05	MBHMG1	8	1.16	8.46	9.62	8.1	82.0	
P5132-06	MBHMG2	9	1.16	8.73	9.89	8.78	87.3	
P5132-07	MBHMH6	10	1.15	8.62	9.77	7.41	72.6	
P5132-08	MBHMH7	11	1.16	8.60	9.76	7.89	78.3	
P5132-09	MBHMH8	12	1.17	8.65	9.82	8.19	81.2	
P5132-10	MBHMH9	13	1.16	8.72	9.88	8.29	81.8	
P5132-11	MBHMJ0	14	1.16	8.48	9.64	8.09	81.7	
P5132-12	MBHMJ1	15	1.16	8.82	9.98	7.9	76.4	
P5132-13	MBHMJ2	16	1.15	8.56	9.71	7.86	78.4	
P5132-14	MBHMJ3	17	1.15	8.39	9.54	7.81	79.4	
P5132-15	MBHMJ4	18	1.15	8.48	9.63	7.81	78.5	
P5132-16	MBHMJ5	19	1.18	8.50	9.68	7.65	76.1	
P5132-17	MBHMJ6	20	1.15	8.83	9.98	7.81	75.4	
P5132-18	MBHMJ6D	21	1.15	8.83	9.98	7.81	75.4	
P5132-19	MBHMJ6S	22	1.15	8.83	9.98	7.81	75.4	
P5132-20	MBHMN4	23	1.19	8.34	9.53	7.99	81.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$