

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133819

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3157

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	12/09/2024	12:50
2	P5145-01		1	14	100.00	>212.0	>212.0	12/09/2024	13:20
3	P5145-01DUP		1	14	100.00	>212.0	>212.0	12/09/2024	13:50
4	P5160-01		1	11	100.00	>212.0	>212.0	12/09/2024	14:20
5	P5170-01		1	14	100.00	>212.0	>212.0	12/09/2024	14:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16133819

WorkList Name : fp-12-07

WorkList ID : 186090

Department : Wet-Chemistry

Date : 12-07-2024 07:56:01

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5145-01	286085	Water	Flash Point	Cool 4 deg C	PSEG03	L51	12/05/2024	1010B
P5160-01	112524	Water	Flash Point	Cool 4 deg C	PSEG03	L61	12/06/2024	1010B
P5170-01	120324-A-B	Water	Flash Point	Cool 4 deg C	PSEG03	L41	12/06/2024	1010B

Date/Time 12/07/2024 12:40
Raw Sample Received by: RM CWCJ
Raw Sample Relinquished by: 28/001

Date/Time 12/09/2024 15:25
Raw Sample Received by: 28/007
Raw Sample Relinquished by: 12/14 CWCJ