



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/10/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:20
In Date: 12/09/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:21
Out Date: 12/10/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133825

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5172-01	MYCZ19	1	1.12	8.72	9.84	9.68	98.2	
P5172-02	MYCZ20	2	1.15	8.48	9.63	9.47	98.1	
P5172-03	MYCZ21	3	1.13	8.80	9.93	9.77	98.2	
P5172-04	MYCZ22	4	1.15	8.54	9.69	9.51	97.9	
P5172-05	MYCZ23	5	1.15	8.59	9.74	9.57	98.0	
P5172-06	MYCZ24	6	1.16	8.41	9.57	9.44	98.5	
P5172-07	MYCZ25	7	1.18	8.77	9.95	9.61	96.1	
P5172-08	MYCZ26	8	1.17	8.61	9.78	9.6	97.9	
P5172-09	MYCZ27	9	1.14	8.55	9.69	9.44	97.1	
P5172-10	MYCZ28	10	1.17	8.57	9.74	9.42	96.3	
P5172-11	MYCZ29	11	1.15	8.43	9.58	9.41	98.0	
P5172-12	MYCZ30	12	1.15	8.53	9.68	9.53	98.2	
P5172-13	MYCZ31	13	1.17	8.81	9.98	9.71	96.9	
P5172-14	MYCZ32	14	1.16	8.80	9.96	9.72	97.3	
P5172-15	MYCZ33	15	1.18	8.43	9.61	9.45	98.1	
P5172-16	MYCZ34	16	1.17	8.41	9.58	9.42	98.1	
P5172-17	MYCZ35	17	1.15	8.59	9.74	9.61	98.5	
P5172-18	MYCZ36	18	1.16	8.69	9.85	9.68	98.0	
P5172-19	MYCZ37	19	1.17	8.59	9.76	9.6	98.1	
P5172-20	MYCZ38	20	1.14	8.54	9.68	9.53	98.2	
P5172-21	MYCZ38D	21	1.14	8.54	9.68	9.53	98.2	
P5172-22	MYCZ38S	22	1.14	8.54	9.68	9.53	98.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$