



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/10/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:35
In Date: 12/09/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:25
Out Date: 12/10/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133826

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5173-01	MYCZ39	1	1.16	8.40	9.56	9.36	97.6	
P5173-02	MYCZ40	2	1.15	8.46	9.61	9.44	98.0	
P5173-03	MYCZ41	3	1.14	8.84	9.98	9.77	97.6	
P5173-04	MYCZ42	4	1.14	8.54	9.68	9.55	98.5	
P5173-05	MYCZ43	5	1.16	8.62	9.78	9.59	97.8	
P5173-06	MYCZ44	6	1.15	8.71	9.86	9.74	98.6	
P5173-07	MYCZ45	7	1.14	8.71	9.85	9.7	98.3	
P5173-08	MYCZ46	8	1.14	8.73	9.87	9.77	98.9	
P5173-09	MYCZ47	9	1.15	8.69	9.84	9.75	99.0	
P5173-10	MYCZ48	10	1.14	8.64	9.78	9.71	99.2	
P5173-11	MYCZ49	11	1.14	8.70	9.84	9.64	97.7	
P5173-12	MYCZ50	12	1.14	8.64	9.78	9.65	98.5	
P5173-13	MYCZ51	13	1.14	8.39	9.53	9.26	96.8	
P5173-14	MYCZ52	14	1.15	8.79	9.94	9.77	98.1	
P5173-15	MYCZ53	15	1.14	8.54	9.68	9.57	98.7	
P5173-16	MYCZ54	16	1.14	8.48	9.62	9.52	98.8	
P5173-17	MYCZ55	17	1.14	8.49	9.63	9.31	96.2	
P5173-18	MYCZ56	18	1.18	8.43	9.61	9.54	99.2	
P5173-19	MYCZ57	19	1.14	8.65	9.79	9.65	98.4	
P5173-20	MYCZ58	20	1.16	8.46	9.62	9.56	99.3	
P5173-21	MYCZ58D	21	1.16	8.46	9.62	9.56	99.3	
P5173-22	MYCZ58S	22	1.16	8.46	9.62	9.56	99.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$