



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/10/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:25
In Date: 12/09/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:30
Out Date: 12/10/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133827

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5177-01	MYCZ59	1	1.18	8.45	9.63	9.57	99.3	
P5177-02	MYCZ60	2	1.15	8.75	9.9	9.52	95.7	
P5177-03	MYCZ61	3	1.16	8.47	9.63	9.47	98.1	
P5177-04	MYCZ62	4	1.16	8.44	9.6	9.2	95.3	
P5177-05	MYCZ63	5	1.18	8.35	9.53	8.99	93.5	
P5177-06	MYCZ64	6	1.17	8.69	9.86	9.78	99.1	
P5177-07	MYCZ65	7	1.17	8.61	9.78	9.51	96.9	
P5177-08	MYCZ66	8	1.16	8.58	9.74	9.46	96.7	
P5177-09	MYCZ67	9	1.18	8.62	9.8	9.66	98.4	
P5177-10	MYCZ68	10	1.16	8.68	9.84	9.74	98.8	
P5177-11	MYD012	11	1.16	8.36	9.52	9.31	97.5	
P5177-12	MYD013	12	1.18	8.79	9.97	9.77	97.7	
P5177-13	MYD014	13	1.17	8.55	9.72	9.51	97.5	
P5177-14	MYD015	14	1.17	8.54	9.71	9.54	98.0	
P5177-15	MYD016	15	1.14	8.41	9.55	9.38	98.0	
P5177-16	MYD017	16	1.18	8.45	9.63	9.54	98.9	
P5177-17	MYD018	17	1.15	8.73	9.88	9.71	98.1	
P5177-18	MYD019	18	1.15	8.60	9.75	9.6	98.3	
P5177-19	MYD020	19	1.15	8.81	9.96	9.7	97.0	
P5177-20	MYD021	20	1.16	8.61	9.77	9.59	97.9	
P5177-21	MYD021D	21	1.16	8.61	9.77	9.59	97.9	
P5177-22	MYD021S	22	1.16	8.61	9.77	9.59	97.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$