



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/10/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:45
In Date: 12/09/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:26
Out Date: 12/10/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133828

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5178-01	MYCZ69	1	1.16	8.50	9.66	9.55	98.7	
P5178-02	MYCZ70	2	1.17	8.71	9.88	9.77	98.7	
P5178-03	MYCZ71	3	1.15	8.48	9.63	9.5	98.5	
P5178-04	MYCZ72	4	1.15	8.54	9.69	9.51	97.9	
P5178-05	MYCZ73	5	1.16	8.52	9.68	9.57	98.7	
P5178-06	MYCZ74	6	1.12	8.40	9.52	9.42	98.8	
P5178-07	MYCZ75	7	1.14	8.68	9.82	9.47	96.0	
P5178-08	MYCZ76	8	1.16	8.61	9.77	9.62	98.3	
P5178-09	MYCZ77	9	1.12	8.40	9.52	9.13	95.4	
P5178-10	MYCZ78	10	1.13	8.40	9.53	9.51	99.8	
P5178-11	MYCZ79	11	1.15	8.56	9.71	9.38	96.1	
P5178-12	MYCZ80	12	1.11	8.55	9.66	9.59	99.2	
P5178-13	MYCZ81	13	1.12	8.52	9.64	9.37	96.8	
P5178-14	MYCZ82	14	1.12	8.62	9.74	9.62	98.6	
P5178-15	MYCZ83	15	1.12	8.86	9.98	9.65	96.3	
P5178-16	MYCZ84	16	1.12	8.72	9.84	9.72	98.6	
P5178-17	MYCZ85	17	1.12	8.45	9.57	9.46	98.7	
P5178-18	MYCZ86	18	1.12	8.40	9.52	9.29	97.3	
P5178-19	MYCZ87	19	1.12	8.75	9.87	9.71	98.2	
P5178-20	MYCZ88	20	1.12	8.63	9.75	9.54	97.6	
P5178-21	MYCZ88D	21	1.12	8.63	9.75	9.54	97.6	
P5178-22	MYCZ88S	22	1.12	8.63	9.75	9.54	97.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$