

LB133829


7470A

INSTRUMENT ID: CVD

Linear

μ Abs.:

41025



A= 0.0000e+000

B= 2.4732e-004

C= -5.9321e-002

Rho= 0.9997465

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/10
0.0	0.000	0.010	0.010	279	0.000	279	0				
0.2	0.200	0.173	-0.027	939	0.0 %	939					-14
2.5	2.500	2.621	0.121	10838	0.0 %	10838					5
5.0	5.000	4.911	-0.089	20098	0.0 %	20098					-2
7.5	7.500	7.398	-0.102	30151	0.0 %	30151					-1
10.0	10.000	10.087	0.087	41025	0.0 %	41025					-1

LB133829 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Dec 2024 10:31:36

Sample ID	Extended ID	μ Abs	Conc	Strd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	279	-	0.000	7470A	PPB	09 Dec 2024 11:15:41	S	Std
0.2 - 1	50-2	939	-	0.200	7470A	PPB	09 Dec 2024 11:21:10	S	Std
2.5 - 1	52-5	10838	-	2.500	7470A	PPB	09 Dec 2024 11:23:27	S	Std
5.0 - 1	52-5	20098	-	5.000	7470A	PPB	09 Dec 2024 11:28:27	S	Std
7.5 - 1	55-5	30151	-	7.500	7470A	PPB	09 Dec 2024 11:33:19	S	Std
10.0 - 1	510	41025	-	10.000	7470A	PPB	09 Dec 2024 11:38:22	S	Std
ICV13 - 1	ICV13	15890	3.8706	-	7470A	PPB	09 Dec 2024 11:46:21	U	SMPL
ICB13 - 1	ICB13	90	-0.0371	-	7470A	PPB	09 Dec 2024 11:48:35	U	SMPL
CCV05 - 1	CCV05	18754	4.5790	-	7470A	PPB	09 Dec 2024 11:50:51	U	SMPL
CCB05 - 1	CCB05	87	-0.0378	-	7470A	PPB	09 Dec 2024 11:53:06	U	SMPL
CRA - 1	CRA	892	0.1613	-	7470A	PPB	09 Dec 2024 11:58:00	U	SMPL
HighStd - 1	HighStd	39391	9.6830	-	7470A	PPB	09 Dec 2024 12:00:15	U	SMPL
ChkStd - 1	ChkStd	26995	6.6172	-	7470A	PPB	09 Dec 2024 12:02:30	U	SMPL
PB165486BL - 1	PBW	437	0.0488	-	7470A	PPB	09 Dec 2024 12:07:24	U	SMPL
PB165486BS - 1	LCSW	13970	3.3958	-	7470A	PPB	09 Dec 2024 12:12:14	U	SMPL
P5095-04 - 1	MH-764	333	0.0230	-	7470A	PPB	09 Dec 2024 12:14:30	U	SMPL
P5096-04 - 1	MH-B	586	0.0856	-	7470A	PPB	09 Dec 2024 12:16:47	U	SMPL
P5096-08 - 1	MH-A	522	0.0698	-	7470A	PPB	09 Dec 2024 12:19:04	U	SMPL
P5100-04 - 1	423	590	0.0866	-	7470A	PPB	09 Dec 2024 12:21:21	U	SMPL
P5103-02 - 1	423	2656	0.5976	-	7470A	PPB	09 Dec 2024 12:23:37	U	SMPL
CCV06 - 1	CCV06	19966	4.8787	-	7470A	PPB	09 Dec 2024 12:25:52	U	SMPL
CCB06 - 1	CCB06	129	-0.0274	-	7470A	PPB	09 Dec 2024 12:28:09	U	SMPL
P5110-01 - 1	ELIZ-COMP-1	2039	0.4450	-	7470A	PPB	09 Dec 2024 12:30:26	U	SMPL
P5110-02 - 1	ELIZ-COMP-2	41	-0.0492	-	7470A	PPB	09 Dec 2024 12:32:41	U	SMPL
P5112-02 - 1	10TH-ST-SOIL	688	0.1108	-	7470A	PPB	09 Dec 2024 12:35:00	U	SMPL
P5117-02 - 1	TAPIAL2-IDW-SOIL-120424-00-T2	301	0.0151	-	7470A	PPB	09 Dec 2024 12:37:15	U	SMPL
P5133-02 - 1	MOO-24-00374	25	-0.0531	-	7470A	PPB	09 Dec 2024 12:39:31	U	SMPL
P5136-02 - 1	COMP-1	338	0.0243	-	7470A	PPB	09 Dec 2024 12:41:47	U	SMPL
P5136-02MS - 1	COMP-1MS	15584	3.7950	-	7470A	PPB	09 Dec 2024 12:46:21	U	SMPL
P5136-02MSD - 1	COMP-1MSD	15878	3.8677	-	7470A	PPB	09 Dec 2024 12:48:38	U	SMPL
PB165487BL - 1	PBW	328	0.0218	-	7470A	PPB	09 Dec 2024 12:50:55	U	SMPL
CCV07 - 1	CCV07	20330	4.9688	-	7470A	PPB	09 Dec 2024 12:53:13	U	SMPL
CCB07 - 1	CCB07	99	-0.0348	-	7470A	PPB	09 Dec 2024 12:55:28	U	SMPL
PB165487BS - 1	LCSW	13521	3.2847	-	7470A	PPB	09 Dec 2024 13:00:24	U	SMPL
P5100-03 - 1	3154	36	-0.0504	-	7470A	PPB	09 Dec 2024 13:02:39	U	SMPL
P5103-01 - 1	422	231	-0.0022	-	7470A	PPB	09 Dec 2024 13:04:55	U	SMPL
P5103-01DUP - 1	422DUP	162	-0.0193	-	7470A	PPB	09 Dec 2024 13:07:11	U	SMPL
P5103-01MS - 1	422MS	9900	2.3892	-	7470A	PPB	09 Dec 2024 13:09:26	U	SMPL
P5103-01MSD - 1	422MSD	10169	2.4557	-	7470A	PPB	09 Dec 2024 13:12:55	U	SMPL
PB165491BL - 1	PBW	-72	-0.0771	-	7470A	PPB	09 Dec 2024 13:15:11	U	SMPL
PB165491BS - 1	LCSW	13338	3.2395	-	7470A	PPB	09 Dec 2024 13:23:31	U	SMPL
P5160-01 - 1	112524	3285	0.7531	-	7470A	PPB	09 Dec 2024 13:25:47	U	SMPL
P5160-01DUP - 1	112524DUP	3173	0.7254	-	7470A	PPB	09 Dec 2024 13:28:03	U	SMPL
CCV08 - 1	CCV08	19373	4.7321	-	7470A	PPB	09 Dec 2024 13:37:32	U	SMPL
CCB08 - 1	CCB08	317	0.0191	-	7470A	PPB	09 Dec 2024 13:48:11	U	SMPL
P5160-01MS - 1	112524MS	8063	1.9348	-	7470A	PPB	09 Dec 2024 13:50:28	U	SMPL
P5160-01MSD - 1	112524MSD	7600	1.8203	-	7470A	PPB	09 Dec 2024 13:52:45	U	SMPL
PB165390TB - 1	PB165390TB	581	0.0844	-	7470A	PPB	09 Dec 2024 14:03:05	U	SMPL
PB165392TB - 1	PB165392TB	430	0.0470	-	7470A	PPB	09 Dec 2024 14:13:13	U	SMPL
P5136-02LX5 - 1		-45	-0.0705	-	7470A	PPB	09 Dec 2024 14:15:31	U	SMPL
P5136-02A - 1		9138	2.2007	-	7470A	PPB	09 Dec 2024 14:17:49	U	SMPL
P5103-01LX5 - 1		-2076	-0.5728	-	7470A	PPB	09 Dec 2024 14:28:43	U	SMPL
P5103-01A - 1		8910	2.1443	-	7470A	PPB	09 Dec 2024 14:31:05	U	SMPL
P5160-01LX5 - 1		484	0.0604	-	7470A	PPB	09 Dec 2024 14:49:32	U	SMPL
P5160-01A - 1		8620	2.0726	-	7470A	PPB	09 Dec 2024 14:51:51	U	SMPL
CCV09 - 1	CCV09	20559	5.0254	-	7470A	PPB	09 Dec 2024 15:27:09	U	SMPL
CCB09 - 1	CCB09	-90	-0.0816	-	7470A	PPB	09 Dec 2024 15:35:05	U	SMPL
P3136-02DUP - 1	COMP-1DUP	36	-0.0504	-	7470A	PPB	09 Dec 2024 15:37:24	U	SMPL
CCV10 - 1	CCV10	21388	5.2304	-	7470A	PPB	09 Dec 2024 15:39:41	U	SMPL
CCB10 - 1	CCB10	-39	-0.0690	-	7470A	PPB	09 Dec 2024 15:47:24	U	SMPL