



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/10/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 15:30
In Date: 12/09/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:42
Out Date: 12/10/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133833

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5182-01	MYCZ9	1	1.16	8.49	9.65	9.48	98.0	
P5182-02	MYCZ00	2	1.15	8.76	9.91	9.75	98.2	
P5182-03	MYCZ01	3	1.15	8.45	9.6	9.47	98.5	
P5182-04	MYCZ02	4	1.18	8.53	9.71	9.57	98.4	
P5182-05	MYCZ03	5	1.17	8.66	9.83	9.72	98.7	
P5182-06	MYCZ04	6	1.18	8.54	9.72	9.62	98.8	
P5182-07	MYCZ05	7	1.18	8.49	9.67	9.52	98.2	
P5182-08	MYCZ06	8	1.17	8.34	9.51	9.46	99.4	
P5182-09	MYCZ07	9	1.17	8.39	9.56	9.31	97.0	
P5182-10	MYCZ08	10	1.17	8.40	9.57	9.38	97.7	
P5182-11	MYCZ09	11	1.17	8.71	9.88	9.75	98.5	
P5182-12	MYCZ10	12	1.18	8.54	9.72	9.47	97.1	
P5182-13	MYCZ11	13	1.15	8.77	9.92	9.8	98.6	
P5182-14	MYCZ12	14	1.16	8.54	9.7	9.47	97.3	
P5182-15	MYCZ13	15	1.17	8.39	9.56	9.55	99.9	
P5182-16	MYCZ14	16	1.16	8.45	9.61	9.45	98.1	
P5182-17	MYCZ15	17	1.18	8.42	9.6	9.47	98.5	
P5182-18	MYCZ16	18	1.17	8.46	9.63	9.48	98.2	
P5182-19	MYCZ17	19	1.17	8.50	9.67	9.44	97.3	
P5182-20	MYCZ18	20	1.15	8.36	9.51	9.38	98.4	
P5182-21	MYCZ18D	21	1.15	8.36	9.51	9.38	98.4	
P5182-22	MYCZ18S	22	1.15	8.36	9.51	9.38	98.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$