

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/10/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:20
In Date: 12/09/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 12/10/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133841

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5184-01	MYCZD9	1	1.15	8.45	9.6	9.43	98.0	
P5184-02	MYCZE0	2	1.15	8.62	9.77	9.68	99.0	
P5184-03	MYCZE1	3	1.14	8.66	9.8	9.67	98.5	
P5184-04	MYCZE2	4	1.16	8.66	9.82	9.57	97.1	
P5184-05	MYCZE3	5	1.16	8.36	9.52	9.32	97.6	
P5184-06	MYCZE4	6	1.15	8.52	9.67	9.58	98.9	
P5184-07	MYCZE5	7	1.16	8.49	9.65	9.44	97.5	
P5184-08	MYCZE6	8	1.16	8.77	9.93	9.77	98.2	
P5184-09	MYCZE7	9	1.16	8.41	9.57	9.38	97.7	
P5184-10	MYCZE8	10	1.15	8.75	9.9	9.62	96.8	
P5184-11	MYCZE9	11	1.18	8.75	9.93	9.68	97.1	
P5184-12	MYCZF0	12	1.15	8.47	9.62	9.47	98.2	
P5184-13	MYCZF1	13	1.16	8.47	9.63	9.44	97.8	
P5184-14	MYCZF2	14	1.15	8.45	9.6	9.45	98.2	
P5184-15	MYCZF3	15	1.16	8.51	9.67	9.46	97.5	
P5184-16	MYCZF4	16	1.17	8.42	9.59	9.42	98.0	
P5184-17	MYCZF5	17	1.16	8.58	9.74	9.55	97.8	
P5184-18	MYCZF6	18	1.16	8.39	9.55	9.4	98.2	
P5184-19	MYCZF7	19	1.18	8.48	9.66	9.51	98.2	
P5184-20	MYCZF8	20	1.18	8.73	9.91	9.77	98.4	
P5184-21	MYCZF8D	21	1.18	8.73	9.91	9.77	98.4	
P5184-22	MYCZF8S	22	1.18	8.73	9.91	9.77	98.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$