

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB133846

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/09/2024	12:25
2	CAL2	1	Water	NA	NA	20.3	7.01	12/09/2024	12:26
3	CAL3	1	Water	NA	NA	20.2	10.03	12/09/2024	12:29
4	ICV	1	Water	NA	NA	20.3	6.99	12/09/2024	12:33
5	CCV1	1	Water	NA	NA	20.3	2.01	12/09/2024	12:35
6	P5216-04	1	Solid	20.02	20	24.5	6.86	12/09/2024	12:40
7	P5216-04DUP	1	Solid	20.03	20	24.6	6.87	12/09/2024	12:41
8	P5216-08	1	Solid	20.02	20	23.7	9.59	12/09/2024	12:47
9	P5216-12	1	Solid	20.03	20	23.6	7.12	12/09/2024	12:55
10	P5216-16	1	Solid	20.02	20	23.8	7.82	12/09/2024	13:00
11	CCV2	1	Water	NA	NA	20.3	12.02	12/09/2024	13:05