

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133855

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/10/2024	08:00
2	CAL2	1	Water	NA	NA	20.2	7.00	12/10/2024	08:02
3	CAL3	1	Water	NA	NA	20.2	10.02	12/10/2024	08:03
4	ICV	1	Water	NA	NA	20.3	7.00	12/10/2024	08:05
5	CCV1	1	Water	NA	NA	20.2	2.02	12/10/2024	08:35
6	P5218-01	1	Water	NA	NA	20.8	7.33	12/10/2024	08:38
7	P5218-01DUP	1	Water	NA	NA	20.9	7.34	12/10/2024	08:39
8	P5220-02	1	Water	NA	NA	22.2	9.02	12/10/2024	08:45
9	P5220-03	1	Water	NA	NA	21.0	10.63	12/10/2024	09:00
10	P5222-02	1	Water	NA	NA	22.1	5.28	12/10/2024	09:10
11	CCV2	1	Water	NA	NA	20.3	12.02	12/10/2024	09:15

WORKLIST(Hardcopy Internal Chain)

VR 133855

WorkList Name : ph p5222

WorkList ID : 186157

Department : Wet-Chemistry

Date : 12-10-2024 07:45:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5218-01	✓ 50385	Water	pH	Cool 4 deg C	PSEG03	L51	12/09/2024	9040C
P5220-02	✓ GAS-AUD-1607-GAS-AUD-1601	Water	pH	Cool 4 deg C	PSEG03	L41	12/09/2024	9040C
P5220-03	✓ GAS-AUD-1610-1611-1612	Water	pH	Cool 4 deg C	PSEG03	L41	12/09/2024	9040C
P5222-02	✓ GAS-TRE-1111	Water	pH	Cool 4 deg C	PSEG03	L51	12/09/2024	9040C

Date/Time 12-10-24 07:50

Raw Sample Received by: 78 WOC

Raw Sample Relinquished by: ✓

Date/Time 12-10-24 12:30

Raw Sample Received by: ✓

Raw Sample Relinquished by: 78 WOC