



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/11/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:50
In Date: 12/10/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:36
Out Date: 12/11/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133859

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5185-01	MYCZF9	1	1.16	8.48	9.64	9.49	98.2	
P5185-02	MYCZG0	2	1.15	8.61	9.76	9.63	98.5	
P5185-03	MYCZG1	3	1.16	8.50	9.66	9.55	98.7	
P5185-04	MYCZG2	4	1.15	8.37	9.52	9.45	99.2	
P5185-05	MYCZG3	5	1.15	8.47	9.62	9.52	98.8	
P5185-06	MYCZG4	6	1.18	8.43	9.61	9.52	98.9	
P5185-07	MYCZG5	7	1.15	8.63	9.78	9.69	99.0	
P5185-08	MYCZG6	8	1.17	8.60	9.77	7.96	79.0	
P5185-09	MYCZG7	9	1.17	8.53	9.7	9.61	98.9	
P5185-10	MYCZG8	10	1.15	8.42	9.57	9.5	99.2	
P5185-11	MYCZM9	11	1.15	8.43	9.58	9.52	99.3	
P5185-12	MYCZN0	12	1.15	8.76	9.91	9.33	93.4	
P5185-13	MYCZN1	13	1.18	8.57	9.75	8.06	80.3	
P5185-14	MYCZN2	14	1.14	8.60	9.74	9.64	98.8	
P5185-15	MYCZN3	15	1.16	8.70	9.86	9.77	99.0	
P5185-16	MYCZN4	16	1.15	8.40	9.55	9.43	98.6	
P5185-17	MYCZN5	17	1.15	8.47	9.62	9.4	97.4	
P5185-18	MYCZN6	18	1.16	8.73	9.89	9.74	98.3	
P5185-19	MYCZN7	19	1.18	8.34	9.52	9.43	98.9	
P5185-20	MYCZN8	20	1.18	8.50	9.68	9.63	99.4	
P5185-21	MYCZN8D	21	1.18	8.50	9.68	9.63	99.4	
P5185-22	MYCZN8S	22	1.18	8.50	9.68	9.63	99.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$