

LB133875


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

47669



A= 0.0000e+000

B= 2.0782e-004

C= -1.6821e-001

Rho= 0.9987661

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	01010
0.0	0.000	-0.083	-0.083	409	0.000	409	0				7
0.2	0.200	0.188	-0.012	1715	0.0 %	1715					6
2.5	2.500	2.584	0.084	13245	0.0 %	13245					3
5.0	5.000	4.934	-0.066	24553	0.0 %	24553					1
7.5	7.500	7.838	0.338	38527	0.0 %	38527					5
10.0	10.000	9.738	-0.262	47669	0.0 %	47669					3

LB133875 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Dec 2024 07:57:16

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	409	-	0.0000	7470A	PPB	11 Dec 2024 08:06:34	S	Std
0.2 - 1	SO.2	1715	-	0.2000	7470A	PPB	11 Dec 2024 08:08:52	S	Std
2.5 - 1	SO.5	13245	-	2.5000	7470A	PPB	11 Dec 2024 08:11:09	S	Std
5.0 - 1	SO	24553	-	5.0000	7470A	PPB	11 Dec 2024 08:13:26	S	Std
7.5 - 1	SO.5	38527	-	7.5000	7470A	PPB	11 Dec 2024 08:18:35	S	Std
10.0 - 1	SO	47669	-	10.0000	7470A	PPB	11 Dec 2024 08:20:50	S	Std
ICV15 - 1	ICV15	18196	3.6132	-	7470A	PPB	11 Dec 2024 08:23:52	U	SMPL
ICB15 - 1	ICB15	48	-0.1582	-	7470A	PPB	11 Dec 2024 08:28:49	U	SMPL
CCV13 - 1	CCV13	24636	4.9515	-	7470A	PPB	11 Dec 2024 08:31:03	U	SMPL
CCB13 - 1	CCB13	32	-0.1616	-	7470A	PPB	11 Dec 2024 08:33:19	U	SMPL
CRA - 1	CRA	1511	0.1458	-	7470A	PPB	11 Dec 2024 08:35:35	U	SMPL
HighStd - 1	HighStd	46270	9.4474	-	7470A	PPB	11 Dec 2024 08:37:51	U	SMPL
ChkStd - 1	ChkStd	36299	7.3753	-	7470A	PPB	11 Dec 2024 08:40:06	U	SMPL
PB165533BL - 1	PBW	375	-0.0903	-	7470A	PPB	11 Dec 2024 08:42:24	U	SMPL
PB165533BS - 1	LCSW	18273	3.6292	-	7470A	PPB	11 Dec 2024 08:44:42	U	SMPL
P5174-02 - 1	ROLL-OFF-COMP	-84	-0.1857	-	7470A	PPB	11 Dec 2024 08:46:59	U	SMPL
P5196-04 - 1	MH-761	220	-0.1225	-	7470A	PPB	11 Dec 2024 08:49:19	U	SMPL
P5216-04 - 1	BP-G-6	504	-0.0635	-	7470A	PPB	11 Dec 2024 08:51:36	U	SMPL
P5216-08 - 1	MH-743	683	-0.0263	-	7470A	PPB	11 Dec 2024 08:53:54	U	SMPL
P5216-12 - 1	MH-744	1185	0.0780	-	7470A	PPB	11 Dec 2024 08:56:08	U	SMPL
CCV14 - 1	CCV14	24327	4.8873	-	7470A	PPB	11 Dec 2024 08:58:23	U	SMPL
CCB14 - 1	CCB14	-6	-0.1695	-	7470A	PPB	11 Dec 2024 09:00:38	U	SMPL
P5216-16 - 1	TP-12	350	-0.0955	-	7470A	PPB	11 Dec 2024 09:02:56	U	SMPL
P5216-16DUP - 1	TP-12DUP	295	-0.1069	-	7470A	PPB	11 Dec 2024 09:05:12	U	SMPL
P5216-16MS - 1	TP-12MS	16488	3.2582	-	7470A	PPB	11 Dec 2024 09:07:27	U	SMPL
P5216-16MSD - 1	TP-12MSD	17880	3.5475	-	7470A	PPB	11 Dec 2024 09:09:43	U	SMPL
PB165470TB - 1	PB165470TB	525	-0.0591	-	7470A	PPB	11 Dec 2024 09:14:40	U	SMPL
P5216-16LX5 - 1		89	-0.1497	-	7470A	PPB	11 Dec 2024 09:16:56	U	SMPL
P5216-16A - 1		18989	3.7780	-	7470A	PPB	11 Dec 2024 09:19:12	U	SMPL
CCV15 - 1	CCV15	24364	4.8950	-	7470A	PPB	11 Dec 2024 09:21:29	U	SMPL
CCB15 - 1	CCB15	322	-0.1013	-	7470A	PPB	11 Dec 2024 09:23:48	U	SMPL