

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133876

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/11/2024	08:05
2	CAL2	1	Water	NA	NA	20.3	7.00	12/11/2024	08:06
3	CAL3	1	Water	NA	NA	20.2	10.02	12/11/2024	08:08
4	ICV	1	Water	NA	NA	20.3	7.01	12/11/2024	08:11
5	CCV1	1	Water	NA	NA	20.3	2.01	12/11/2024	08:15
6	P5243-01	1	Water	NA	NA	20.6	5.83	12/11/2024	08:25
7	P5243-01DUP	1	Water	NA	NA	20.7	5.84	12/11/2024	08:26
8	P5243-03	1	Water	NA	NA	21.6	5.37	12/11/2024	08:33
9	P5243-04	1	Water	NA	NA	21.4	5.29	12/11/2024	08:37
10	P5243-05	1	Water	NA	NA	21.8	5.29	12/11/2024	08:38
11	P5243-06	1	Water	NA	NA	21.7	5.19	12/11/2024	08:44
12	CCV2	1	Water	NA	NA	20.2	12.02	12/11/2024	08:45

WORKLIST(Hardcopy Internal Chain)

43133816

WorkList Name : ph p5243

WorkList ID : 186200

Department : Wet-Chemistry

Date : 12-11-2024 07:57:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5243-01	1203	Water	pH	Cool 4 deg C	PSEG03	L51	12/10/2024	9040C
P5243-03	1127-A	Water	pH	Cool 4 deg C	PSEG03	L51	12/10/2024	9040C
P5243-04	1127-B	Water	pH	Cool 4 deg C	PSEG03	L51	12/10/2024	9040C
P5243-05	1127-C	Water	pH	Cool 4 deg C	PSEG03	L51	12/10/2024	9040C
P5243-06	1127-D	Water	pH	Cool 4 deg C	PSEG03	L51	12/10/2024	9040C

Date/Time 12-11-24 08:00

Raw Sample Received by: JB WDC

Raw Sample Relinquished by: JF SN

Date/Time 12-11-24 10:00

Raw Sample Received by: JF SN

Raw Sample Relinquished by: JF WDC