

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133896

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3157

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	12/11/2024	10:00
2	P5161-01		1	6	41.00	105.8	106.1	12/11/2024	10:30
3	P5218-01		1	11	100.00	>212.0	>212.0	12/11/2024	11:00
4	P5220-02		1	14	100.00	>212.0	>212.0	12/11/2024	11:30
5	P5220-03		1	14	100.00	>212.0	>212.0	12/11/2024	12:00
6	P5222-02		1	12	100.00	>212.0	>212.0	12/11/2024	12:30
7	P5222-02DUP		1	12	100.00	>212.0	>212.0	12/11/2024	13:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB133896

WorkList Name : FP-12-09

WorkList ID : 186153

Department : Wet-Chemistry

Date : 12-09-2024 14:46:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5161-01	125	Water	Flash Point	Cool 4 deg C	PSEG03	L51	12/06/2024	1010B
P5218-01	50385	Water	Flash Point	Cool 4 deg C	PSEG03	L51	12/09/2024	1010B
P5220-02	GAS-AUD-1607-GAS-AUD-1601	Water	Flash Point	Cool 4 deg C	PSEG03	L41	12/09/2024	1010B
P5220-03	GAS-AUD-1610-1611-1612	Water	Flash Point	Cool 4 deg C	PSEG03	L41	12/09/2024	1010B
P5222-02	GAS-TRE-1111	Water	Flash Point	Cool 4 deg C	PSEG03	L51	12/09/2024	1010B

Date/Time 12/11/2024 09:50
Raw Sample Received by: RY
Raw Sample Relinquished by:

Date/Time 12/11/2024 13:40
Raw Sample Received by:
Raw Sample Relinquished by: