

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133899

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3157

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	12/11/2024	15:10
2	P5161-02		1	4	7.00	44.6	44.9	12/11/2024	15:40
3	P5161-02DUP		1	4	7.00	44.6	44.9	12/11/2024	16:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP12-11

WorkList ID : 186250

Department : Wet-Chemistry

Date : 12-11-2024 09:43:21

65133899

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5161-02	125	Solid	Flash Point	Cool 4 deg C	PSEG03	L51	12/06/2024	1010B

Date/Time 12/11/2024 15:00
 Raw Sample Received by: R14 CWG
 Raw Sample Relinquished by: JH CWG

Date/Time 12/11/2024 16:50
 Raw Sample Received by: JH CWG
 Raw Sample Relinquished by: RM CWG