

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133903

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/12/2024	09:00
2	CAL2	1	Water	NA	NA	20.2	7.00	12/12/2024	09:01
3	CAL3	1	Water	NA	NA	20.2	10.02	12/12/2024	09:05
4	ICV	1	Water	NA	NA	20.3	7.00	12/12/2024	09:11
5	CCV1	1	Water	NA	NA	20.3	2.01	12/12/2024	09:15
6	P5236-01	1	Solid	20.02	20	23.3	4.96	12/12/2024	09:25
7	P5236-01DUP	1	Solid	20.03	20	23.4	4.97	12/12/2024	09:26
8	CCV2	1	Water	NA	NA	20.3	12.02	12/12/2024	09:30

WORKLIST(Hardcopy Internal Chain)

✓ 133903

WorkList Name : ph p5236

WorkList ID : 186266

Department : Wet-Chemistry

Date : 12-12-2024 08:42:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5236-01	✓ TAPIAL3-SB03D-3-120724-00-1	Solid	pH	Cool 4 deg C	WEST04	L51	12/07/2024	9045D

Date/Time 12-12-24 08:50
Raw Sample Received by: ph wll
Raw Sample Relinquished by: CSM

Date/Time 12-12-24 11:00
Raw Sample Received by: CSM
Raw Sample Relinquished by: ph wll