


LB133907

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.0178e-004 slope

C= -1.4699e-002 y-intercept

Rho= 0.9999285

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/n
0.00	0.000	0.055	0.055	347	0.000	347	0				11
0.05	0.050					442	0				5
0.20	0.200	0.209	0.009	1109	0.0 %	1109					-2
2.50	2.500	2.459	-0.041	12259	0.0 %	12259					-1
5.00	5.000	4.953	-0.047	24618	0.0 %	24618					0
7.50	7.500	7.465	-0.035	37069	0.0 %	37069					1
10.0	10.000	10.059	0.059	49921	0.0 %	49921					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	347	-		0 SFAM01.1	PPB	12/12/2024 10:54	Std
	0.2 S01	1109	-		0.2 SFAM01.1	PPB	12/12/2024 10:56	Std
	2.5 S02	12259	-		2.5 SFAM01.1	PPB	12/12/2024 11:00	Std
	5 S03	24618	-		5 SFAM01.1	PPB	12/12/2024 11:03	Std
	7.5 S04	37069	-		7.5 SFAM01.1	PPB	12/12/2024 11:05	Std
	10 S05	49921	-		10 SFAM01.1	PPB	12/12/2024 11:12	Std
ICV017	ICV017	20985	4.2197 -		SFAM01.1	PPB	12/12/2024 11:17	SMPL
ICB017	ICB017	282	0.0422 -		SFAM01.1	PPB	12/12/2024 11:19	SMPL
CCV020	CCV020	24276	4.8838 -		SFAM01.1	PPB	12/12/2024 11:22	SMPL
CCB020	CCB020	327	0.0513 -		SFAM01.1	PPB	12/12/2024 11:24	SMPL
PB165593BL	PBS593	463	0.0787 -		SFAM01.1	PPB	12/12/2024 11:26	SMPL
P5067-01	ME28M0	2109	0.4109 -		SFAM01.1	PPB	12/12/2024 11:28	SMPL
P5067-02	ME28M1	856	0.158 -		SFAM01.1	PPB	12/12/2024 11:31	SMPL
P5067-03	ME28M2	675	0.1215 -		SFAM01.1	PPB	12/12/2024 11:33	SMPL
P5067-04	ME28M3	620	0.1104 -		SFAM01.1	PPB	12/12/2024 11:35	SMPL
P5067-05	ME28M4	641	0.1146 -		SFAM01.1	PPB	12/12/2024 11:38	SMPL
P5067-06	ME28M5	612	0.1088 -		SFAM01.1	PPB	12/12/2024 11:40	SMPL
P5067-07	ME28M6	456	0.0773 -		SFAM01.1	PPB	12/12/2024 11:42	SMPL
P5067-08	ME28M7	2048	0.3986 -		SFAM01.1	PPB	12/12/2024 11:44	SMPL
P5067-09	ME28M8	1772	0.3429 -		SFAM01.1	PPB	12/12/2024 11:47	SMPL
P5067-10	ME28M9	1455	0.2789 -		SFAM01.1	PPB	12/12/2024 11:49	SMPL
P5067-11	ME28N0	1835	0.3556 -		SFAM01.1	PPB	12/12/2024 11:51	SMPL
P5067-12	ME28N1	846	0.156 -		SFAM01.1	PPB	12/12/2024 11:53	SMPL
P5067-13	ME28N2	1140	0.2153 -		SFAM01.1	PPB	12/12/2024 11:56	SMPL
P5067-14	ME28N3	744	0.1354 -		SFAM01.1	PPB	12/12/2024 11:58	SMPL
P5067-15	ME28N4	3194	0.6298 -		SFAM01.1	PPB	12/12/2024 12:00	SMPL
P5067-16	ME28N5	1149	0.2172 -		SFAM01.1	PPB	12/12/2024 12:02	SMPL
P5067-17	ME28N6	1443	0.2765 -		SFAM01.1	PPB	12/12/2024 12:05	SMPL
P5067-18	ME28N7	546	0.0955 -		SFAM01.1	PPB	12/12/2024 12:07	SMPL
P5067-19	ME28N8	25	-0.0097 -		SFAM01.1	PPB	12/12/2024 12:09	SMPL
CCV021	CCV021	24687	4.9667 -		SFAM01.1	PPB	12/12/2024 12:12	SMPL
CCB021	CCB021	393	0.0646 -		SFAM01.1	PPB	12/12/2024 12:14	SMPL
P5067-20	ME28N8D	223	0.0303 -		SFAM01.1	PPB	12/12/2024 12:19	SMPL
P5067-21	ME28N8S	12807	2.5696 -		SFAM01.1	PPB	12/12/2024 12:21	SMPL
PB165594BL	PBS594	190	0.0236 -		SFAM01.1	PPB	12/12/2024 12:23	SMPL
P5154-01	ME28N9	1291	0.2458 -		SFAM01.1	PPB	12/12/2024 12:25	SMPL
P5154-02	ME28P0	1254	0.2383 -		SFAM01.1	PPB	12/12/2024 12:28	SMPL
P5154-03	ME28P1	636	0.1136 -		SFAM01.1	PPB	12/12/2024 12:30	SMPL
P5154-04	ME28P2	1059	0.199 -		SFAM01.1	PPB	12/12/2024 12:32	SMPL
P5154-05	ME28P3	129438	26.1038 -		SFAM01.1	PPB	12/12/2024 12:35	SMPL
P5154-06	ME28P4	794	0.1455 -		SFAM01.1	PPB	12/12/2024 12:37	SMPL
P5154-07	ME28P5	1608	0.3098 -		SFAM01.1	PPB	12/12/2024 12:39	SMPL
P5154-08	ME28P6	861	0.159 -		SFAM01.1	PPB	12/12/2024 12:42	SMPL
P5154-09	ME28P7	490	0.0842 -		SFAM01.1	PPB	12/12/2024 12:44	SMPL
P5154-10	ME28P8	1776	0.3437 -		SFAM01.1	PPB	12/12/2024 12:46	SMPL
P5154-11	ME28P9	1900	0.3687 -		SFAM01.1	PPB	12/12/2024 12:48	SMPL

INSTRUMENT ID : CV1

P5154-12	ME28Q0	1505	0.289 -	SFAM01.1	PPB	12/12/2024 12:51	SMPL
P5154-13	ME28Q1	2011	0.3911 -	SFAM01.1	PPB	12/12/2024 12:53	SMPL
P5154-14	ME28Q2	1634	0.315 -	SFAM01.1	PPB	12/12/2024 12:55	SMPL
P5154-15	ME28Q3	6884	1.3744 -	SFAM01.1	PPB	12/12/2024 12:57	SMPL
P5154-16	ME28Q4	6140	1.2243 -	SFAM01.1	PPB	12/12/2024 13:00	SMPL
P5154-17	ME28Q4D	6603	1.3177 -	SFAM01.1	PPB	12/12/2024 13:02	SMPL
CCV022	CCV022	23367	4.7004 -	SFAM01.1	PPB	12/12/2024 13:04	SMPL
CCB022	CCB022	117	0.0089 -	SFAM01.1	PPB	12/12/2024 13:06	SMPL
P5154-18	ME28Q4S	17680	3.5528 -	SFAM01.1	PPB	12/12/2024 13:09	SMPL
P5154-19	ME28Q5	2094	0.4078 -	SFAM01.1	PPB	12/12/2024 13:11	SMPL
P5154-20	ME28R3	4485	0.8903 -	SFAM01.1	PPB	12/12/2024 13:13	SMPL
P5154-21	ME28R4	4278	0.8485 -	SFAM01.1	PPB	12/12/2024 13:16	SMPL
PB165595BL	PBS595	175	0.0206 -	SFAM01.1	PPB	12/12/2024 13:18	SMPL
P5233-01	ME28R5	1999	0.3887 -	SFAM01.1	PPB	12/12/2024 13:20	SMPL
P5233-02	ME28R6	1990	0.3869 -	SFAM01.1	PPB	12/12/2024 13:22	SMPL
P5233-03	ME28R7	647	0.1159 -	SFAM01.1	PPB	12/12/2024 13:25	SMPL
P5233-04	ME28R8	-8	-0.0163 -	SFAM01.1	PPB	12/12/2024 13:27	SMPL
P5233-05	ME28R9	380	0.062 -	SFAM01.1	PPB	12/12/2024 13:29	SMPL
P5233-06	ME28S0	5091	1.0126 -	SFAM01.1	PPB	12/12/2024 13:31	SMPL
P5233-07	ME28S1	1106	0.2085 -	SFAM01.1	PPB	12/12/2024 13:34	SMPL
P5233-08	ME28S2	2882	0.5668 -	SFAM01.1	PPB	12/12/2024 13:36	SMPL
P5233-09	ME28S3	1272	0.242 -	SFAM01.1	PPB	12/12/2024 13:38	SMPL
P5233-10	ME28S4	1446	0.2771 -	SFAM01.1	PPB	12/12/2024 13:40	SMPL
P5233-11	ME28S5	1388	0.2654 -	SFAM01.1	PPB	12/12/2024 13:43	SMPL
P5233-12	ME28S6	1367	0.2611 -	SFAM01.1	PPB	12/12/2024 13:45	SMPL
P5233-13	ME28S7	1288	0.2452 -	SFAM01.1	PPB	12/12/2024 13:47	SMPL
P5233-14	ME28S8	1540	0.296 -	SFAM01.1	PPB	12/12/2024 13:50	SMPL
P5233-15	ME28S9	998	0.1867 -	SFAM01.1	PPB	12/12/2024 13:52	SMPL
CCV023	CCV023	23447	4.7165 -	SFAM01.1	PPB	12/12/2024 13:54	SMPL
CCB023	CCB023	137	0.0129 -	SFAM01.1	PPB	12/12/2024 13:56	SMPL
P5233-16	ME28T0	1148	0.2169 -	SFAM01.1	PPB	12/12/2024 13:59	SMPL
P5233-17	ME28T1	8183	1.6365 -	SFAM01.1	PPB	12/12/2024 14:01	SMPL
P5233-18	ME28T2	907	0.1683 -	SFAM01.1	PPB	12/12/2024 14:03	SMPL
P5233-19	ME28T3	4470	0.8873 -	SFAM01.1	PPB	12/12/2024 14:05	SMPL
P5233-20	ME28S7D	4171	0.8269 -	SFAM01.1	PPB	12/12/2024 14:08	SMPL
P5233-21	ME28S7S	13789	2.7677 -	SFAM01.1	PPB	12/12/2024 14:13	SMPL
P5154-05DLX5	ME28P3	24602	4.9496 -	SFAM01.1	PPB	12/12/2024 14:15	SMPL
CCV024	CCV024	25284	5.0872 -	SFAM01.1	PPB	12/12/2024 14:20	SMPL
CCB024	CCB024	207	0.0271 -	SFAM01.1	PPB	12/12/2024 14:23	SMPL