

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133908

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3157

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	12/12/2024	13:45
2	P5243-01		1	12	100.00	>212.0	>212.0	12/12/2024	14:15
3	P5243-01DUP		1	12	100.00	>212.0	>212.0	12/12/2024	14:45
4	P5243-02		1	14	100.00	>212.0	>212.0	12/12/2024	15:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6133908

WorkList Name : fp 12-12

WorkList ID : 186257

Department : Wet-Chemistry

Date : 12-12-2024 08:18:57

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5243-01	1203	Water	Flash Point	Cool 4 deg C	PSEG03	L51	12/10/2024	1010B
P5243-02	1127-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	L51	12/10/2024	1010B

Date/Time 12/12/2024 13:35
Raw Sample Received by: RM CWG
Raw Sample Relinquished by: JH (CWG)

Date/Time 12/12/2024 15:50
Raw Sample Received by: JH (CWG)
Raw Sample Relinquished by: RM CWG