

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB133926

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/12/2024	11:15
2	CAL2	1	Water	NA	NA	20.3	7.00	12/12/2024	11:16
3	CAL3	1	Water	NA	NA	20.3	10.02	12/12/2024	11:20
4	ICV	1	Water	NA	NA	20.3	7.01	12/12/2024	11:22
5	CCV1	1	Water	NA	NA	20.3	2.01	12/12/2024	11:25
6	P5239-04	1	Solid	20.02	20	23.4	6.84	12/12/2024	11:35
7	P5239-04DUP	1	Solid	20.03	20	23.5	6.85	12/12/2024	11:36
8	CCV2	1	Water	NA	NA	20.3	12.02	12/12/2024	11:40