


LB133933

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.040	0.040	-57	0.000	-57	0				-2
0.2	0.200	0.204	0.004	711	0.0 %	711					9
2.5	2.500	2.735	0.235	12511	0.0 %	12511					-8
5.0	5.000	4.604	-0.396	21229	0.0 %	21229					0
7.5	7.500	7.503	0.003	34752	0.0 %	34752					1
10.0	10.000	10.114	0.114	46928	0.0 %	46928					

LB133933 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Dec 2024 09:18:07

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-57	-	0.0000	7470A	PPB	13 Dec 2024	10:05:13	S Std
0.2 - 1	50.2	711	-	0.2000	7470A	PPB	13 Dec 2024	10:07:29	S Std
2.5 - 1	52.5	12511	-	2.5000	7470A	PPB	13 Dec 2024	10:09:46	S Std
5.0 - 1	55	21229	-	5.0000	7470A	PPB	13 Dec 2024	10:15:53	S Std
7.5 - 1	57.5	34752	-	7.5000	7470A	PPB	13 Dec 2024	10:18:10	S Std
10.0 - 1	59.5	46928	-	10.0000	7470A	PPB	13 Dec 2024	10:20:28	S Std
ICV18 - 1	ICV18	18712	4.0641	-	7470A	PPB	13 Dec 2024	10:26:13	U SMPL
ICB18 - 1	ICB18	-260	-0.0038	-	7470A	PPB	13 Dec 2024	10:28:29	U SMPL
CCV25 - 1	CCV25	25076	5.4287	-	7470A	PPB	13 Dec 2024	10:30:46	U SMPL
CCB25 - 1	CCB25	-264	-0.0047	-	7470A	PPB	13 Dec 2024	10:33:01	U SMPL
CRA - 1	CRA	857	0.2357	-	7470A	PPB	13 Dec 2024	10:35:19	U SMPL
HighStd - 1	HighStd	46981	10.1255	-	7470A	PPB	13 Dec 2024	10:37:34	U SMPL
ChkStd - 1	ChkStd	32415	7.0023	-	7470A	PPB	13 Dec 2024	10:39:49	U SMPL
PB165598BL - 1	PBW	-251	-0.0019	-	7470A	PPB	13 Dec 2024	10:42:07	U SMPL
PB165598BS - 1	LCSW	17627	3.8315	-	7470A	PPB	13 Dec 2024	10:47:05	U SMPL
P5239-04 - 1	WC-1	-156	0.0185	-	7470A	PPB	13 Dec 2024	10:49:21	U SMPL
P5239-04DUP - 1	WC-1DUP	-16	0.0485	-	7470A	PPB	13 Dec 2024	10:51:40	U SMPL
P5239-04MS - 1	WC-1MS	17726	3.8527	-	7470A	PPB	13 Dec 2024	10:56:39	U SMPL
P5239-04MSD - 1	WC-1MSD	18547	4.0288	-	7470A	PPB	13 Dec 2024	10:58:56	U SMPL
PB165599BL - 1	PBW	-24	0.0468	-	7470A	PPB	13 Dec 2024	11:01:13	U SMPL
CCV26 - 1	CCV26	24687	5.3453	-	7470A	PPB	13 Dec 2024	11:03:30	U SMPL
CCB26 - 1	CCB26	-258	-0.0034	-	7470A	PPB	13 Dec 2024	11:05:46	U SMPL
PB165599BS - 1	LCSW	16630	3.6177	-	7470A	PPB	13 Dec 2024	11:11:29	U SMPL
P5224-02 - 1	NP-WS-001	-176	0.0142	-	7470A	PPB	13 Dec 2024	11:13:44	U SMPL
P5224-02DUP - 1	NP-WS-001DUP	73	0.0678	-	7470A	PPB	13 Dec 2024	11:16:01	U SMPL
P5224-02MS - 1	NP-WS-001MS	10122	2.2223	-	7470A	PPB	13 Dec 2024	11:18:17	U SMPL
P5224-02MSD - 1	NP-WS-001MSD	9946	2.1845	-	7470A	PPB	13 Dec 2024	11:20:32	U SMPL
PB165616BL - 1	PBW	-129	0.0243	-	7470A	PPB	13 Dec 2024	11:22:49	U SMPL
PB165616BS - 1	LCSW	16879	3.6711	-	7470A	PPB	13 Dec 2024	11:27:44	U SMPL
P4368-09 - 1	MDL-WATER-03-QT4-2024	-9	0.0500	-	7470A	PPB	13 Dec 2024	11:30:01	U SMPL
P4368-07 - 1	LOD-MDL-WATER	35	0.0594	-	7470A	PPB	13 Dec 2024	11:46:56	U SMPL
P4368-07RE - 1	LOD-MDL-WATERRE	579	0.1761	-	7470A	PPB	13 Dec 2024	11:49:11	U SMPL
CCV27 - 1	CCV27	24182	5.2370	-	7470A	PPB	13 Dec 2024	11:51:26	U SMPL
CCB27 - 1	CCB27	-257	-0.0032	-	7470A	PPB	13 Dec 2024	11:53:42	U SMPL
P5220-03 - 1	GAS-AUD-1610-1611-1612	230	0.1012	-	7470A	PPB	13 Dec 2024	11:55:59	U SMPL
P5220-03DUP - 1	GAS-AUD-1610-1611-1612DUP	209	0.0967	-	7470A	PPB	13 Dec 2024	11:58:14	U SMPL
P5220-03MS - 1	GAS-AUD-1610-1611-1612MS	9516	2.0923	-	7470A	PPB	13 Dec 2024	12:00:30	U SMPL
P5220-03MSD - 1	GAS-AUD-1610-1611-1612MSD	8098	1.7883	-	7470A	PPB	13 Dec 2024	12:02:46	U SMPL
P5230-01 - 1	266	-257	-0.0032	-	7470A	PPB	13 Dec 2024	12:05:03	U SMPL
P5231-01 - 1	270	89	0.0710	-	7470A	PPB	13 Dec 2024	12:07:21	U SMPL
PB165558TB - 1	PB165558TB	-31	0.0453	-	7470A	PPB	13 Dec 2024	12:09:37	U SMPL
PB165561TB - 1	PB165561TB	120	0.0777	-	7470A	PPB	13 Dec 2024	12:11:54	U SMPL
CCV28 - 1	CCV28	22321	4.8380	-	7470A	PPB	13 Dec 2024	12:14:11	U SMPL
CCB28 - 1	CCB28	-221	0.0045	-	7470A	PPB	13 Dec 2024	12:16:27	U SMPL
P5239-04LX5 - 1		92	0.0717	-	7470A	PPB	13 Dec 2024	12:18:44	U SMPL
P5239-04A - 1		15488	3.3728	-	7470A	PPB	13 Dec 2024	12:20:59	U SMPL
P5224-02LX5 - 1		-103	0.0298	-	7470A	PPB	13 Dec 2024	12:23:15	U SMPL
P5224-02A - 1		15652	3.4080	-	7470A	PPB	13 Dec 2024	12:25:32	U SMPL
P5220-03LX5 - 1		-104	0.0296	-	7470A	PPB	13 Dec 2024	12:27:47	U SMPL
P5220-03A - 1		14993	3.2667	-	7470A	PPB	13 Dec 2024	12:39:59	U SMPL
CCV29 - 1	CCV29	23290	5.0457	-	7470A	PPB	13 Dec 2024	12:42:14	U SMPL
CCB29 - 1	CCB29	-218	0.0052	-	7470A	PPB	13 Dec 2024	12:44:31	U SMPL