



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/17/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:25
In Date: 12/16/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:45
Out Date: 12/17/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133956

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5226-01	E2AN9	1	1.15	8.40	9.55	8.27	84.8	
P5226-02	E2AP0	2	1.15	8.83	9.98	9.14	90.5	
P5226-03	E2AP1	3	1.14	8.77	9.91	9.07	90.4	
P5226-04	E2AP2	4	1.12	8.76	9.88	8.33	82.3	
P5226-05	E2AP3	5	1.19	8.50	9.69	9.1	93.1	
P5226-06	E2AP4	6	1.14	8.84	9.98	9.14	90.5	
P5226-07	E2AP5	7	1.12	8.77	9.89	8.38	82.8	
P5226-08	E2AP6	8	1.12	8.48	9.6	8.99	92.8	
P5226-09	E2AP7	9	1.15	8.43	9.58	8.79	90.6	
P5226-10	VHBLK001	10	1.00	1.00	2.00	2.00	100.0	vhblk
P5226-11	E2AN6	11	1.13	8.80	9.93	8.37	82.3	
P5226-12	E2AQ4	12	1.15	8.82	9.97	8.68	85.4	
P5226-13	E2AQ5	13	1.15	8.40	9.55	8.72	90.1	
P5226-14	E2AQ6	14	1.16	8.47	9.63	8.51	86.8	
P5226-15	E2AQ7	15	1.12	8.70	9.82	8.00	79.1	
P5226-16	E2AQ8	16	1.17	8.70	9.87	8.71	86.7	
P5226-17	E2AR3	17	1.16	8.69	9.85	8.29	82.0	
P5226-18	E2AR4	18	1.14	8.85	9.99	8.9	87.7	
P5226-19	E2AQ4	19	1.15	8.82	9.97	8.68	85.4	
P5226-20	E2AQ6	20	1.16	8.47	9.63	8.51	86.8	
P5226-21	E2AQ7	21	1.12	8.70	9.82	8.00	79.1	
P5226-22	E2AQ8	22	1.17	8.70	9.87	8.71	86.7	
P5226-23	E2AR3	23	1.16	8.69	9.85	8.29	82.0	
P5226-24	E2AR4	24	1.14	8.85	9.99	8.9	87.7	
P5226-25	E2AR4MS	25	1.14	8.85	9.99	8.9	87.7	
P5226-26	E2AR4MSD	26	1.14	8.85	9.99	8.9	87.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$